

LB105954

=====

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by :

Instrument ID : Konelab

10/28/2019 9:02


Test Total CN

Accepted 10/28/2019 9:02

Factor 1179
Bias 0.002

Coeff. of det. 0.999967

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.001	-0.7490	0.0000	
2	5.0PPBCN	0.006	4.6536	5.0000	
3	10PPBCN	0.010	9.8429	10.0000	
4	50PPBCN	0.044	49.9832	50.0000	
5	100PPBCN	0.087	100.2238	100.0000	
6	250PPBCN	0.216	252.1638	250.0000	
7	500PPBCN	0.425	498.8816	500.0000	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : Coel

Instrument ID : Konelab

10/28/2019 10:09

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	97.983 ✓	0.0	0.085	
ICB1	-0.436 ✓	0.0	0.001	
CCV1	245.706 ✓	0.0	0.210	
CCB1	-0.230 ✓	0.0	0.002	
PB124313BL	-0.332 ✓	0.0	0.002	
PB124313BS	97.642 ✓	0.0	0.085	
LOWPB124313	10.144 ✓	0.0	0.010	
HIGHPB124313	494.568 ✓	0.0	0.421	
K5153-05MS	32.766 ✓	0.0	0.030	
K5153-05MSD	33.444 ✓	0.0	0.030	
CCV2	250.759 ✓	0.0	0.215	
CCB2	0.092 ✓	0.0	0.002	
K5543-01	0.691 ✓	0.0	0.002	
K5544-01	-0.415 ✓	0.0	0.002	
K5545-01	-0.464 ✓	0.0	0.001	
K5153-05	-0.061 ✓	0.0	0.002	
K5153-05DUP	0.634 ✓	0.0	0.002	
CCV3	262.786 ✓	0.0	0.225	
CCB3	-0.167 ✓	0.0	0.002	

N	19
Mean	80.269
SD	136.6094
CV%	170.19

ee
10/28/19

Aquakem v. 7.2AQ1

Results from time period:

Mon Oct 28 08:47:25 2019

Mon Oct 28 10:06:00 2019

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	-0.749	µg/l	10/28/2019 8:47:25	
5.0PPBCN	A	Total CN	P	4.6536	µg/l	10/28/2019 8:47:26	
10PPBCN	A	Total CN	P	9.8429	µg/l	10/28/2019 8:47:27	
50PPBCN	A	Total CN	P	49.9832	µg/l	10/28/2019 8:47:28	
100PPBCN	A	Total CN	P	100.2238	µg/l	10/28/2019 8:47:29	
250PPBCN	A	Total CN	P	252.1638	µg/l	10/28/2019 8:47:30	
500PPBCN	A	Total CN	P	498.8816	µg/l	10/28/2019 8:47:31	
ICV1	S	Total CN	P	97.9829	µg/l	10/28/2019 9:33:41	
ICB1	S	Total CN	P	-0.4356	µg/l	10/28/2019 9:33:42	
CCV1	S	Total CN	P	245.7065	µg/l	10/28/2019 9:33:45	
CCB1	S	Total CN	P	-0.2298	µg/l	10/28/2019 9:33:46	
PB124313BL	S	Total CN	P	-0.3319	µg/l	10/28/2019 9:33:48	
PB124313BS	S	Total CN	P	97.642	µg/l	10/28/2019 9:33:49	
LOWPB124313	S	Total CN	P	10.1444	µg/l	10/28/2019 9:33:50	
HIGHPB124313	S	Total CN	P	494.5682	µg/l	10/28/2019 9:41:13	
K5153-05MS	S	Total CN	P	32.7656	µg/l	10/28/2019 9:41:16	
K5153-05MSD	S	Total CN	P	33.444	µg/l	10/28/2019 9:41:18	
CCV2	S	Total CN	P	250.7594	µg/l	10/28/2019 9:41:20	
CCB2	S	Total CN	P	0.0923	µg/l	10/28/2019 9:41:21	
K5543-01	S	Total CN	P	0.6906	µg/l	10/28/2019 9:41:22	
K5544-01	S	Total CN	P	-0.4145	µg/l	10/28/2019 9:41:23	
K5545-01	S	Total CN	P	-0.4641	µg/l	10/28/2019 9:45:32	
K5153-05	S	Total CN	P	-0.0613	µg/l	10/28/2019 10:05:57	
K5153-05DUP	S	Total CN	P	0.6341	µg/l	10/28/2019 10:05:58	
CCV3	S	Total CN	P	262.7861	µg/l	10/28/2019 10:05:59	
CCB3	S	Total CN	P	-0.1671	µg/l	10/28/2019 10:06:00	