



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: AMANDEEP

Run Number: LB105960

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	10/28/2019	11:32
2	K5569-02		1	9	100.00	>212.0	>212.0	10/28/2019	11:52
3	K5569-02DUP		1	9	100.00	>212.0	>212.0	10/28/2019	12:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 105960

WorkList Name : fp102819 WorkList ID : 132661 Department : Wet-Chemistry Date : 10-28-2019 09:18:27

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/30/2019	Water	K5569-02	Flash Point	Cool 4 deg C	JACO05	IDW-DECON-WATER-102519	10/25/2019	1010A	

Date/Time 10/28/19 11:00
 Received by: KS
 Relinquished by: JP

Date/Time 10/28/19 12:25
 Received by: JP
 Relinquished by: KS