



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB105961

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	10/28/2019	11:50
2	CAL2	1	Water	NA	NA	20.1	7.01	10/28/2019	11:51
3	CAL3	1	Water	NA	NA	20.1	10.04	10/28/2019	11:52
4	ICV	1	Water	NA	NA	20.1	6.99	10/28/2019	12:00
5	CCV1	1	Water	NA	NA	20.2	2.01	10/28/2019	12:02
6	K5580-02	1	Water	NA	NA	18.4	7.51	10/28/2019	12:05
7	K5580-02DUP	1	Water	NA	NA	18.5	7.52	10/28/2019	12:06
8	K5580-03	1	Water	NA	NA	18.5	6.48	10/28/2019	12:07
9	CCV2	1	Water	NA	NA	20.1	12.02	10/28/2019	12:10

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K5580 WorkList ID : 132674 Department : Wet-Chemistry Date : 10-28-2019 11:44:29

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/25/2019	Water	K5580-02- <u>1</u>	pH	Cool 4 deg C	TETR06	E22	TT-007-IDWGW-20191025	10/25/2019	9040C
10/25/2019	Water	K5580-03- <u>1</u>	pH	Cool 4 deg C	TETR06	E22	TT-008-IDWGW-20191025	10/25/2019	9040C

Date/Time 10-28-19 11:45
 Received by: Jr
 Relinquished by: CS

Date/Time 10-28-19 12-35
 Received by: CS
 Relinquished by: Jr