

LB105970


7471B

INSTRUMENT ID: CV1

Linear ▼

 μ Abs.:

54909



A= 0.0000e+000

B= 1.8231e-004

C= -3.7670e-002

Rho= 0.9999525

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.017	-0.017	115	0.000	115				
0.2	0.200	0.173	-0.027	1153	0.0 %	1153				
2.5	2.500	2.517	0.017	14010	0.0 %	14010				
5.0	5.000	5.073	0.073	28034	0.0 %	28034				
7.5	7.500	7.481	-0.019	41243	0.0 %	41243				
10.0	10.000	9.973	-0.027	54909	0.0 %	54909				

LB105970 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 28 Oct 2019 09:04:53

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	115	-	0.00007471B	PPB		28 Oct 2019 09:16:14	S	Std
0.2 - 1	50-2	1153	-	0.20007471B	PPB		28 Oct 2019 09:18:39	S	Std
2.5 - 1	52-5	14010	-	2.50007471B	PPB		28 Oct 2019 09:20:57	S	Std
5.0 - 1	55-0	28034	-	5.00007471B	PPB		28 Oct 2019 09:23:15	S	Std
7.5 - 1	57-5	41243	-	7.50007471B	PPB		28 Oct 2019 09:25:32	S	Std
10.0 - 1	510-0	54909	-	10.00007471B	PPB		28 Oct 2019 09:27:49	S	Std
ICV76 - 1	ICV76	22697	4.1003	-7471B	PPB		28 Oct 2019 09:30:43	U	SMPL
ICB76 - 1	ICB76	121	-0.0156	-7471B	PPB		28 Oct 2019 09:32:59	U	SMPL
CCV25 - 1	CCV25	28388	5.1378	-7471B	PPB		28 Oct 2019 09:35:15	U	SMPL
CCB25 - 1	CCB25	103	-0.0189	-7471B	PPB		28 Oct 2019 09:37:31	U	SMPL
CRA - 1	CRA	1174	0.1764	-7471B	PPB		28 Oct 2019 09:39:48	U	SMPL
HighStd - 1	HighStd	55521	10.0845	-7471B	PPB		28 Oct 2019 09:42:04	U	SMPL
ChkStd - 1	ChkStd	38496	6.9806	-7471B	PPB		28 Oct 2019 09:44:20	U	SMPL
CCV36 - 1	CCV36	29081	5.2642	-7471B	PPB		28 Oct 2019 14:38:58	U	SMPL
CCB36 - 1	CCB36	231	0.0044	-7471B	PPB		28 Oct 2019 14:41:14	U	SMPL
PB124346BL - 1	PBS	253	0.0085	-7471B	PPB		28 Oct 2019 14:43:30	U	SMPL
PB124346BS - 1	LCSS	21232	3.8332	-7471B	PPB		28 Oct 2019 14:47:15	U	SMPL
K5540-01 - 1	154-82-(0-2)	44322	8.0428	-7471B	PPB		28 Oct 2019 14:49:31	U	SMPL
K5540-03 - 1	155-40-(0-1.1)	125307	22.8074	-7471B	PPB		28 Oct 2019 14:51:47	U	SMPL
K5540-05 - 1	155-98-(0-1.1)	299594	54.5820	-7471B	PPB		28 Oct 2019 14:54:04	U	SMPL
K5540-07 - 1	156-56-(0-1.3)	16190	2.9140	-7471B	PPB		28 Oct 2019 14:56:22	U	SMPL
K5540-09 - 1	157-14-(0-1.6)	65105	11.8318	-7471B	PPB		28 Oct 2019 14:58:43	U	SMPL
CCV37 - 1	CCV37	29535	5.3469	-7471B	PPB		28 Oct 2019 15:00:59	U	SMPL
CCB37 - 1	CCB37	192	-0.0027	-7471B	PPB		28 Oct 2019 15:03:18	U	SMPL
K5540-11 - 1	157-72-(0-1.1)	85689	15.5845	-7471B	PPB		28 Oct 2019 15:05:36	U	SMPL
K5540-13 - 1	158-30-(0-1)	129873	23.6398	-7471B	PPB		28 Oct 2019 15:07:54	U	SMPL
K5540-15 - 1	158-88-(0-1)	27477	4.9717	-7471B	PPB		28 Oct 2019 15:10:11	U	SMPL
K5540-17 - 1	159-46-(0-1)	8765	1.5603	-7471B	PPB		28 Oct 2019 15:12:29	U	SMPL
K5540-19 - 1	160-04-(0-1)	8015	1.4236	-7471B	PPB		28 Oct 2019 15:14:46	U	SMPL
K5541-01 - 1	154-24-(0-2)	61401	11.1565	-7471B	PPB		28 Oct 2019 15:17:02	U	SMPL
K5541-01DUP - 1	154-24-(0-2)DUP	59191	10.7536	-7471B	PPB		28 Oct 2019 15:19:18	U	SMPL
K5541-01MS - 1	154-24-(0-2)MS	87671	15.9458	-7471B	PPB		28 Oct 2019 15:27:11	U	SMPL
K5541-01MSD - 1	154-24-(0-2)MSD	91744	16.6884	-7471B	PPB		28 Oct 2019 15:29:27	U	SMPL
K5541-03 - 1	153-63-(0-2)	22845	4.1273	-7471B	PPB		28 Oct 2019 15:31:46	U	SMPL
CCV38 - 1	CCV38	29254	5.2957	-7471B	PPB		28 Oct 2019 15:34:04	U	SMPL
CCB38 - 1	CCB38	-44	-0.0457	-7471B	PPB		28 Oct 2019 15:36:21	U	SMPL
K5541-05 - 1	153-08-(0-1.1)	3734	0.6431	-7471B	PPB		28 Oct 2019 15:38:37	U	SMPL
K5541-07 - 1	152-50-(0-1)	4871	0.8504	-7471B	PPB		28 Oct 2019 15:40:54	U	SMPL
K5541-09 - 1	51-00-(0-1)	66037	12.0017	-7471B	PPB		28 Oct 2019 15:43:10	U	SMPL
K5541-11 - 1	17-20-(0-2.5)	8766	1.5605	-7471B	PPB		28 Oct 2019 15:45:26	U	SMPL
K5541-13 - 1	15-70-(0-2)	17037	3.0684	-7471B	PPB		28 Oct 2019 15:47:43	U	SMPL
K5541-15 - 1	18-12-(0-2.5)	7598	1.3475	-7471B	PPB		28 Oct 2019 15:49:59	U	SMPL
K5541-17 - 1	18-48-(0-2.5)	12239	2.1937	-7471B	PPB		28 Oct 2019 15:52:15	U	SMPL
K5541-19 - 1	52-77-(0-1.8)	6000	1.0562	-7471B	PPB		28 Oct 2019 15:54:31	U	SMPL
PB124347BL - 1	PBS	141	-0.0120	-7471B	PPB		28 Oct 2019 15:56:48	U	SMPL
PB124347BS - 1	LCSS	21561	3.8932	-7471B	PPB		28 Oct 2019 15:59:04	U	SMPL
CCV39 - 1	CCV39	28332	5.1276	-7471B	PPB		28 Oct 2019 16:01:21	U	SMPL
CCB39 - 1	CCB39	136	-0.0129	-7471B	PPB		28 Oct 2019 16:03:38	U	SMPL
K5542-01 - 1	54-54-(0-1.2)	8076	1.4347	-7471B	PPB		28 Oct 2019 16:05:55	U	SMPL
K5542-01DUP - 1	54-54-(0-1.2)DUP	8028	1.4259	-7471B	PPB		28 Oct 2019 16:08:13	U	SMPL
K5542-01MS - 1	54-54-(0-1.2)MS	28855	5.2230	-7471B	PPB		28 Oct 2019 16:10:30	U	SMPL
K5542-01MSD - 1	54-54-(0-1.2)MSD	27790	5.0288	-7471B	PPB		28 Oct 2019 16:12:46	U	SMPL
K5542-03 - 1	199-48-(0-1.2)	31697	5.7411	-7471B	PPB		28 Oct 2019 16:15:02	U	SMPL
K5542-05 - 1	200-06-(0-1.2)	75735	13.7698	-7471B	PPB		28 Oct 2019 16:17:19	U	SMPL
K5542-07 - 1	198-90-(0-2)	44731	8.1173	-7471B	PPB		28 Oct 2019 16:19:35	U	SMPL
K5542-09 - 1	200-64-(0-1)	84745	15.4124	-7471B	PPB		28 Oct 2019 16:21:53	U	SMPL
K5542-11 - 1	201-22-(0-1)	36430	6.6040	-7471B	PPB		28 Oct 2019 16:24:11	U	SMPL
K5542-13 - 1	201-80-(0-1)	110041	20.0242	-7471B	PPB		28 Oct 2019 16:26:29	U	SMPL
CCV40 - 1	CCV40	27933	5.0549	-7471B	PPB		28 Oct 2019 16:28:47	U	SMPL
CCB40 - 1	CCB40	78	-0.0234	-7471B	PPB		28 Oct 2019 16:31:07	U	SMPL
K5542-15 - 1	202-38-(0-1.5)	58238	10.5798	-7471B	PPB		28 Oct 2019 16:33:24	U	SMPL
K5542-17 - 1	608-60-(0-1.5)	27748	5.0211	-7471B	PPB		28 Oct 2019 16:35:42	U	SMPL
K5541-01LX5 - 1		11458	2.0513	-7471B	PPB		28 Oct 2019 16:38:00	U	SMPL
K5541-01A - 1		86969	15.8179	-7471B	PPB		28 Oct 2019 16:40:17	U	SMPL
K5542-01LX5 - 1		1454	0.2274	-7471B	PPB		28 Oct 2019 16:42:33	U	SMPL
K5542-01A - 1		28114	5.0879	-7471B	PPB		28 Oct 2019 16:44:51	U	SMPL
K5540-03DLX5 - 1		25276	4.5705	-7471B	PPB		28 Oct 2019 16:47:08	U	SMPL
K5540-05DLX10 - 1		32436	5.8758	-7471B	PPB		28 Oct 2019 16:49:25	U	SMPL
K5540-09DLX2 - 1		33706	6.1074	-7471B	PPB		28 Oct 2019 16:51:42	U	SMPL
K5540-11DLX5 - 1		17270	3.1109	-7471B	PPB		28 Oct 2019 16:53:58	U	SMPL
CCV41 - 1	CCV41	29117	5.2707	-7471B	PPB		28 Oct 2019 16:56:16	U	SMPL
CCB41 - 1	CCB41	149	-0.0105	-7471B	PPB		28 Oct 2019 16:58:33	U	SMPL
K5540-13DLX5 - 1		26807	4.8496	-7471B	PPB		28 Oct 2019 17:00:52	U	SMPL
K5541-01DLX2 - 1		31401	5.6871	-7471B	PPB		28 Oct 2019 17:03:09	U	SMPL
K5541-01DUPDLX2 - 1		30112	5.4521	-7471B	PPB		28 Oct 2019 17:05:26	U	SMPL
K5541-01MSDLX2 - 1		42972	7.7967	-7471B	PPB		28 Oct 2019 17:07:42	U	SMPL
K5541-01MSDDLX2 - 1		48620	8.8264	-7471B	PPB		28 Oct 2019 17:09:59	U	SMPL
K5541-01DL2X10 - 1		5889	1.0360	-7471B	PPB		28 Oct 2019 17:12:16	U	SMPL
K5541-01ADLX2 - 1		45751	8.3033	-7471B	PPB		28 Oct 2019 17:14:34	U	SMPL
K5541-09DLX4 - 1		17315	3.1191	-7471B	PPB		28 Oct 2019 17:16:51	U	SMPL
K5542-05DLX4 - 1		39	0.0306	-7471B	PPB		28 Oct 2019 17:19:08	U	SMPL

RS 10-28-19

LB105970 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 28 Oct 2019 09:04:53

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
K5542-09DLX4-1		103	0.0025		-7471B	PPB	28 Oct 2019 17:21:23	U	SMPL
CCV42-1K5542-05DLX4	CCV42	19203	3.4633		-7471B	PPB	28 Oct 2019 17:23:39	U	SMPL
CCB42-1K5542-09DLX4	CCB42	21374	3.8591		-7471B	PPB	28 Oct 2019 17:25:57	U	SMPL
K5542-13DLX5-1		16995	3.0607		-7471B	PPB	28 Oct 2019 17:28:14	U	SMPL
K5542-15DLX2-1		29558	5.3511		-7471B	PPB	28 Oct 2019 17:30:32	U	SMPL
CCV43-1CCV42	CCV43CCV42	28409	5.1416		-7471B	PPB	28 Oct 2019 17:32:49	U	SMPL
CCB43-1CCB42	CCB43CCB42	29	-0.0324		-7471B	PPB	28 Oct 2019 17:35:09	U	SMPL

AS 10-30-15

NS 10-28-15