



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB105972

Slope : 98.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	10/28/2019	15:30
2	CAL2	1	Water	NA	NA	20.1	7.01	10/28/2019	15:31
3	CAL3	1	Water	NA	NA	20.1	10.03	10/28/2019	15:33
4	ICV	1	Water	NA	NA	20.1	7.00	10/28/2019	15:35
5	CCV1	1	Water	NA	NA	20.1	2.02	10/28/2019	15:36
6	K5585-01	1	Water	NA	NA	20.3	5.37	10/28/2019	15:38
7	K5585-01DUP	1	Water	NA	NA	20.4	5.38	10/28/2019	15:39
8	CCV2	1	Water	NA	NA	20.1	12.02	10/28/2019	15:40

LB 105972

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K5585 WorkList ID : 132687 Department : Wet-Chemistry Date : 10-28-2019 15:20:01

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/21/2019	Water	K5585-01-17	pH	Cool 4 deg C	STER11	E22	POOL-WATER	10/21/2019	SM 4500-H B

Date/Time 10-28-19 15:25
 Received by: Jg
 Relinquished by: Jg

Date/Time 10-28-19 16:25
 Received by: Jg
 Relinquished by: Jg