

TOTAL SOLIDS - SM2540B

SUPERVISOR: apatel

ANALYST: AMANDEEP

Date: 10/29/2019

OVEN IN TEMP1: 104 °C

OVEN OUT TEMP1: 105 °C

DATE & TIME IN1: 10/28/2019 15:30 DATE & TIME OUT1: 10/29/2019 09:00

OVEN IN TEMP2: 103 °C

OVEN OUT TEMP2: 104 °C

DATE & TIME IN2: 10/29/2019 09:50 DATE & TIME OUT2: 10/29/2019 10:50

Run Number: LB105980

BalanceID: WC SC-3

OvenID: WC Oven#1

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB105980BL	LB105980BL	91.7256	91.7256	100	91.7256	91.7256	91.7256	0	0
2	K5569-02	IDW-DECON-WATER-102519	92.7551	92.7551	100	93.4438	93.4438	93.4438	0.6887	6887
3	K5569-02DUP	IDW-DECON-WATER-102519DUP	95.0742	95.0742	100	95.7803	95.7803	95.7803	0.7061	7061

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

LB105980

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS102819-2 WorkList ID : 132677 Department : Wet-Chemistry Date : 10-28-2019 11:57:27

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/30/2019	Water	K5569-02	TS	Cool 4 deg C	JACO05	D61	IDW-DECON-WATER-102519	10/25/2019	SM2540 B

Date/Time 10-28-19 15:00
Received by: AK
Relinquished by: CR

Date/Time 10-28-19 16:15
Received by: CR
Relinquished by: AK