

TOTAL VOLATILE SOLIDS 160.4

TEMP1 IN: 104 °C 10/28/2019 15:30 **TEMP1 OUT:** 105 °C 10/29/2019 09:00
TEMP2 IN: 103 °C 10/29/2019 09:50 **TEMP2 OUT:** 104 °C 10/29/2019 10:50
TEMP3 IN: 540 °C 10/29/2019 11:30 **TEMP3 OUT:** 560 °C 10/29/2019 12:30
TEMP4 IN: 560 °C 10/29/2019 13:00 **TEMP4 OUT:** 560 °C 10/29/2019 14:05

Run Number: LB105988

SUPERVISOR: apatel

ANALYST: AMANDEEP

BalanceID: WC SC-3

OvenID: WC Oven#1

Dish #	Lab ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish + SampleWt Drying @103-@105°C (g)	Final Dish + SampleWt Drying @103-@105°C (g)	Dish + Samplewt Drying @550 (±50) °C (g)	Final Dish + Samplewt Drying @550 (±50) °C (g)	Weight Diff (g)	Result (mg/L)
1	LB105988BL	91.7256	91.7256	100	91.7256	91.7256	91.7256	91.7256	0	0
2	K5569-02	92.7551	92.7551	100	93.4438	93.4438	93.3296	93.3296	0.1142	1142
3	K5569-02DUP	95.0742	95.0742	100	95.7803	95.7803	95.6632	95.6632	0.1171	1171

A = Sample Volume (ml)

B = Final Dish + Samplewt Drying @550 (±50) °C (g)

C = Final Dish + SampleWt Drying @103-@105°C (g)

D = Weight (g)

E = Final Empty Dish Weight (g)

F = Final Dish + SampleWt Drying @103-@105°C (g)

Weight D = C - B

Result % = $\frac{D}{A} \star 1000000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TVS102819-2 WorkList ID : 132678 Department : Wet-Chemistry Date : 10-28-2019 11:57:31

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/30/2019	Water	K5569-02	TVS	Cool 4 deg C	JACO05	D61	IDW-DECON-WATER-102519	10/25/2019	160.4

Date/Time 10-28-19 15:00
 Received by: AK
 Relinquished by: AK

Date/Time 10-28-19 16:05
 Received by: AK
 Relinquished by: AK