



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB105992

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP79452
Turbidity Calibration std, 0NTU	WP79446
Turbidity Calibration std, 20NTU	WP79448
Turbidity Calibration std, 40NTU	WP79447
Turbidity Calibration std, 1NTU	WP79451
Turbidity Calibration std, 10NTU	WP79449
Turbidity Calibration std, 5NTU	WP79450
10 NTU Standard 500 ml	W2489

Intercept: -0.038

Slope: 0.9586

Regression: 0.9996

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.190	10/29/2019	13:06
2	CAL2	1	1	0.992	10/29/2019	13:12
3	CAL3	5	1	4.880	10/29/2019	13:18
4	CAL4	10	1	9.560	10/29/2019	13:25
5	CAL5	20	1	18.300	10/29/2019	13:29
6	CAL6	40	1	38.700	10/29/2019	13:35



Analytical Summary Report

Analysis Method: SM2130 B
Parameter: Turbidity
Run Number: LB105992

ANALYST: AMANDEEP
SUPERVISOR REVIEW BY: apatel

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	8.780	9.199	10/29/2019	13:39
2	ICB		1	0.196	0.244	10/29/2019	13:43
3	CCV1	10	1	9.550	10.002	10/29/2019	13:49
4	CCB1		1	0.199	0.247	10/29/2019	13:52
5	LB105992BL		1	0.187	0.235	10/29/2019	13:55
6	K5598-01		1	0.449	0.508	10/29/2019	13:58
7	K5598-01DUP		1	0.440	0.499	10/29/2019	14:01
8	K5598-02		1	0.497	0.558	10/29/2019	14:06
9	K5598-03		1	0.510	0.572	10/29/2019	14:10
10	K5598-04		1	18.400	19.234	10/29/2019	14:15
11	CCV2	10	1	9.600	10.054	10/29/2019	14:17
12	CCB2		1	0.190	0.238	10/29/2019	14:20

WORKLIST(Hardcopy Internal Chain)

LB105992

WorkList Name : turb102919 WorkList ID : 132721 Department : Wet-Chemistry Date : 10-29-2019 12:39:06

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/30/2019	Water	K5598-01	H Turbidity	Cool 4 deg C	LOCK01	D61	MW-1	10/28/2019	SM2130 B
10/30/2019	Water	K5598-02	D Turbidity	Cool 4 deg C	LOCK01	D61	MW-3	10/28/2019	SM2130 B
10/30/2019	Water	K5598-03	H Turbidity	Cool 4 deg C	LOCK01	D61	MW-4	10/28/2019	SM2130 B
10/30/2019	Water	K5598-04	E Turbidity	Cool 4 deg C	LOCK01	D61	MW-2	10/28/2019	SM2130 B

Date/Time 10-29-19 12:45
 Received by: AK
 Relinquished by: CP

Date/Time 10-29-19 14:25
 Received by: CP
 Relinquished by: AK