



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 10/30/2019

OVENTEMP IN Celsius(°C): 107
Time IN: 17:10
In Date: 10/29/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 10/29/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB106004

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
K5148-13	OR-03-102919	1	1.14	9.88	9.13	91.4	
K5148-14	OR-03-102919	2	1.14	9.73	9.09	92.5	
K5148-15	OR-03-102919-1	3	1.13	9.58	8.58	88.2	
K5148-16	OR-03-102919-2	4	1.14	9.91	8.59	84.9	
K5149-07	EO-01-102819	5	1.14	9.73	8.78	88.9	
K5589-01	391024116	6	1.14	9.77	3.16	23.4	SLUDGE SAMPLE
K5599-01	Z3-20A	7	1.15	9.76	9.26	94.2	
K5599-02	Z3-20A	8	1.15	9.76	9.26	94.2	
K5599-03	Z3-20B	9	1.15	9.57	8.9	92.0	
K5599-04	Z3-20B	10	1.15	9.57	8.9	92.0	
K5599-05	Z3-21A	11	1.14	9.63	9.22	95.2	
K5599-06	Z3-21A	12	1.14	9.63	9.22	95.2	
K5599-07	Z3-21B	13	1.14	9.55	9.1	94.6	
K5599-08	Z3-21B	14	1.14	9.55	9.1	94.6	
K5599-09	Z3-22A	15	1.14	9.77	9.36	95.2	
K5599-10	Z3-22A	16	1.14	9.77	9.36	95.2	
K5599-11	Z3-22B	17	1.13	9.54	9.05	94.2	
K5599-12	Z3-22B	18	1.13	9.54	9.05	94.2	
K5599-13	Z3-23A	19	1.13	9.83	9.38	94.8	
K5599-14	Z3-23A	20	1.13	9.83	9.38	94.8	
K5599-15	Z3-23B	21	1.14	9.73	8.73	88.4	
K5599-16	Z3-23B	22	1.14	9.73	8.73	88.4	
K5601-01	SOIL-CUTTINGS	23	1.15	9.77	8.00	79.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$