



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB106005

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	10/29/2019	17:20
2	CAL2	1	Water	NA	NA	20.1	6.99	10/29/2019	17:21
3	CAL3	1	Water	NA	NA	20.1	10.03	10/29/2019	17:22
4	ICV	1	Water	NA	NA	20.1	7.01	10/29/2019	17:25
5	CCV1	1	Water	NA	NA	20.1	2.02	10/29/2019	17:26
6	K5148-17	1	Solid	20.03	20	21.7	8.56	10/29/2019	17:29
7	K5148-17DUP	1	Solid	20.04	20	21.8	8.57	10/29/2019	17:30
8	CCV2	1	Water	NA	NA	20.1	12.02	10/29/2019	17:32

28106005

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY K5999

WorkList ID : 132727

Department : Wet-Chemistry

Date : 10-29-2019 14:50:27

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/29/2019	Solid	K5148-17 - 17	Corrosivity	Cool 4 deg C	PSEG05		OR-03-102919	10/29/2019	9045D

Date/Time 10.29.19 16:10
 Received by: JO
 Relinquished by: PC

Date/Time 10.29.19 18:00
 Received by: PC
 Relinquished by: JO