


LB106011

7471B

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 1.6190e-004

C= -1.5424e-002

Rho= 0.9986863

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.006	0.006	135	0.000	135				
0.2	0.200	0.173	-0.027	1165	0.0 %	1165				
2.5	2.500	2.656	0.156	16502	0.0 %	16502				
5.0	5.000	5.039	0.039	31219	0.0 %	31219				
7.5	7.500	7.119	-0.381	44067	0.0 %	44067				
10.0	10.000	10.206	0.206	63138	0.0 %	63138				

LB106011 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 30 Oct 2019 09:11:28

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	135	-	0.0000	7471B	PPB	30 Oct 2019 09:35:56	S	Std
0.2 - 1	50.2	1165	-	0.2000	7471B	PPB	30 Oct 2019 09:38:14	S	Std
2.5 - 1	52.5	16502	-	2.5000	7471B	PPB	30 Oct 2019 09:40:34	S	Std
5.0 - 1	55.0	31219	-	5.0000	7471B	PPB	30 Oct 2019 09:45:43	S	Std
7.5 - 1	57.5	44067	-	7.5000	7471B	PPB	30 Oct 2019 09:48:01	S	Std
10.0 - 1	510.0	63138	-	10.0000	7471B	PPB	30 Oct 2019 09:54:01	S	Std
ICV80 - 1	ICV80	23846	3.8452	-	7471B	PPB	30 Oct 2019 09:56:55	U	SMPL
ICB80 - 1	ICB80	117	0.0035	-	7471B	PPB	30 Oct 2019 09:59:12	U	SMPL
CCV50 - 1	CCV50	31183	5.0330	-	7471B	PPB	30 Oct 2019 10:01:27	U	SMPL
CCB50 - 1	CCB50	-9	-0.0169	-	7471B	PPB	30 Oct 2019 10:03:43	U	SMPL
CRA - 1	CRA	1214	0.1811	-	7471B	PPB	30 Oct 2019 10:05:59	U	SMPL
HighStd - 1	HighStd	62088	10.0364	-	7471B	PPB	30 Oct 2019 10:08:15	U	SMPL
ChkStd - 1	ChkStd	42427	6.8534	-	7471B	PPB	30 Oct 2019 10:10:31	U	SMPL
PB124408BL - 1	PBS	149	0.0087	-	7471B	PPB	30 Oct 2019 10:12:49	U	SMPL
PB124408BS - 1	LCSS	20934	3.3737	-	7471B	PPB	30 Oct 2019 10:15:07	U	SMPL
K5148-13 - 1	OR-03-102919	1018	0.1494	-	7471B	PPB	30 Oct 2019 10:17:25	U	SMPL
K5149-07 - 1	EO-01-102819	2688	0.4198	-	7471B	PPB	30 Oct 2019 10:21:15	U	SMPL
K5149-07DUP - 1	EO-01-102819DUP	2602	0.4058	-	7471B	PPB	30 Oct 2019 10:23:33	U	SMPL
K5149-07MS - 1	EO-01-102819MS	24082	3.8834	-	7471B	PPB	30 Oct 2019 10:25:50	U	SMPL
K5149-07MSD - 1	EO-01-102819MSD	23608	3.8066	-	7471B	PPB	30 Oct 2019 10:28:06	U	SMPL
CCV51 - 1	CCV51	31551	5.0926	-	7471B	PPB	30 Oct 2019 10:30:22	U	SMPL
CCB51 - 1	CCB51	0	-0.0154	-	7471B	PPB	30 Oct 2019 10:32:38	U	SMPL
K5542-13 - 1	201-80-(0-1)	100844	16.3109	-	7471B	PPB	30 Oct 2019 10:34:55	U	SMPL
K5542-15 - 1	202-38-(0-1.5)	55115	8.9075	-	7471B	PPB	30 Oct 2019 10:37:11	U	SMPL
K5599-01 - 1	Z3-20A	146	0.0082	-	7471B	PPB	30 Oct 2019 10:39:29	U	SMPL
K5599-03 - 1	Z3-20B	132	0.0059	-	7471B	PPB	30 Oct 2019 10:43:16	U	SMPL
K5599-05 - 1	Z3-21A	541	0.0722	-	7471B	PPB	30 Oct 2019 10:47:02	U	SMPL
K5599-07 - 1	Z3-21B	371	0.0446	-	7471B	PPB	30 Oct 2019 10:49:20	U	SMPL
K5599-09 - 1	Z3-22A	348	0.0409	-	7471B	PPB	30 Oct 2019 10:51:37	U	SMPL
K5599-11 - 1	Z3-22B	317	0.0359	-	7471B	PPB	30 Oct 2019 10:53:55	U	SMPL
K5599-13 - 1	Z3-23A	635	0.0874	-	7471B	PPB	30 Oct 2019 10:56:13	U	SMPL
K5599-15 - 1	Z3-23B	229	0.0217	-	7471B	PPB	30 Oct 2019 10:58:29	U	SMPL
CCV52 - 1	CCV52	31140	5.0261	-	7471B	PPB	30 Oct 2019 11:00:45	U	SMPL
CCB52 - 1	CCB52	-18	-0.0183	-	7471B	PPB	30 Oct 2019 11:03:02	U	SMPL
K5601-01 - 1	SOIL-CUTTINGS	1267	0.1897	-	7471B	PPB	30 Oct 2019 11:05:18	U	SMPL
K5149-07LX5 - 1		512	0.0675	-	7471B	PPB	30 Oct 2019 11:07:34	U	SMPL
K5149-07A - 1		23746	3.8290	-	7471B	PPB	30 Oct 2019 11:09:51	U	SMPL
K5542-13DLX5 - 1		20361	3.2810	-	7471B	PPB	30 Oct 2019 11:12:07	U	SMPL
K5542-15DLX2 - 1		28287	4.5642	-	7471B	PPB	30 Oct 2019 11:14:24	U	SMPL
CCV53 - 1	CCV53	31398	5.0678	-	7471B	PPB	30 Oct 2019 11:16:41	U	SMPL
CCB53 - 1	CCB53	51	-0.0072	-	7471B	PPB	30 Oct 2019 11:23:12	U	SMPL