

ISM02.4


LB106018

INSTRUMENT ID: C V1

Linear

 μ Abs.:

41127



A= 0.0000e+000

B= 2.4120e-004

C= -9.0929e-003

Rho= 0.9998701

Accept=Accepted

slope
y intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.004	0.004	56	0.000	56					—
0.05	0.050										
0.20	0.200	0.175	-0.025	762	0.0 %	762					-12
2.50	2.500	2.470	-0.030	10278	0.0 %	10278					-1
5.00	5.000	5.040	0.040	20933	0.0 %	20933					-1
7.50	7.500	7.600	0.100	31549	0.0 %	31549					-1
10.0	10.000	9.911	-0.089	41127	0.0 %	41127					-1

RS 10-30-19

LB106018
INSTRUMENT ID : CV1

Sample ID	Extended II μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	
0 S0.0	56	-	0	ISM02.4	PPB	10/30/2019 12:29	Std	
0.2 S0.2	762	-	0.2	ISM02.4	PPB	10/30/2019 12:31	Std	
2.5 S2.5	10278	-	2.5	ISM02.4	PPB	10/30/2019 12:34	Std	
5 S5.0	20933	-	5	ISM02.4	PPB	10/30/2019 12:36	Std	
7.5 S7.5	31549	-	7.5	ISM02.4	PPB	10/30/2019 12:38	Std	
10 S10.0	41127	-	10	ISM02.4	PPB	10/30/2019 12:40	Std	
ICV	ICV	16768	4.0353	-	ISM02.4	PPB	10/30/2019 12:44	SMPL
ICB	ICB	28	-0.0023	-	ISM02.4	PPB	10/30/2019 12:46	SMPL
CCV054	CCV054	20870	5.0247	-	ISM02.4	PPB	10/30/2019 12:48	SMPL
CCB054	CCB054	-12	-0.012	-	ISM02.4	PPB	10/30/2019 12:51	SMPL
PB124416E	PBW416	20	-0.0043	-	ISM02.4	PPB	10/30/2019 12:53	SMPL
K5389-01	MH0GM0	90	0.0126	-	ISM02.4	PPB	10/30/2019 12:55	SMPL
K5389-02	MH0GM1	208	0.0411	-	ISM02.4	PPB	10/30/2019 12:57	SMPL
K5389-03	MH0GM3	252	0.0517	-	ISM02.4	PPB	10/30/2019 13:00	SMPL
K5389-04	MH0GM4	661	0.1503	-	ISM02.4	PPB	10/30/2019 13:02	SMPL
K5389-05	MH0GM5	1709	0.4031	-	ISM02.4	PPB	10/30/2019 13:04	SMPL
K5389-06	MH0GM8	868	0.2003	-	ISM02.4	PPB	10/30/2019 13:06	SMPL
K5389-07	MH0GM9	109	0.0172	-	ISM02.4	PPB	10/30/2019 13:09	SMPL
K5389-08	MH0GN0	114	0.0184	-	ISM02.4	PPB	10/30/2019 13:11	SMPL
K5389-09	MH0GN2	894	0.2065	-	ISM02.4	PPB	10/30/2019 13:13	SMPL
K5389-10	MH0GN3	71	0.008	-	ISM02.4	PPB	10/30/2019 13:16	SMPL
K5389-11	MH0GN5	196	0.0382	-	ISM02.4	PPB	10/30/2019 13:18	SMPL
K5389-12	MH0GO0	77	0.0095	-	ISM02.4	PPB	10/30/2019 13:20	SMPL
K5389-13	MH0GO2	37	-0.0002	-	ISM02.4	PPB	10/30/2019 13:22	SMPL
K5389-14	MH0GO5	71	0.008	-	ISM02.4	PPB	10/30/2019 13:25	SMPL
K5389-15	MH0GO6	58	0.0049	-	ISM02.4	PPB	10/30/2019 13:27	SMPL
K5389-16	MH0GO7	94	0.0136	-	ISM02.4	PPB	10/30/2019 13:29	SMPL
K5389-17	MH0GO8	249	0.051	-	ISM02.4	PPB	10/30/2019 13:32	SMPL
K5389-18	MH0GQ0	312	0.0662	-	ISM02.4	PPB	10/30/2019 13:34	SMPL
K5389-19	MH0GR3	153	0.0278	-	ISM02.4	PPB	10/30/2019 13:36	SMPL
CCV055	CCV055	20927	5.0384	-	ISM02.4	PPB	10/30/2019 13:38	SMPL
CCB055	CCB055	40	0.0006	-	ISM02.4	PPB	10/30/2019 13:41	SMPL
K5389-20	MH0GR8	106	0.0165	-	ISM02.4	PPB	10/30/2019 13:43	SMPL
K5389-21	MH0GO0D	87	0.0119	-	ISM02.4	PPB	10/30/2019 13:45	SMPL
K5389-22	MH0GO0S	5000	1.1969	-	ISM02.4	PPB	10/30/2019 13:47	SMPL
PB124417E	PBW417	26	-0.0028	-	ISM02.4	PPB	10/30/2019 13:50	SMPL
K5390-01	MH0GR9	29	-0.0021	-	ISM02.4	PPB	10/30/2019 13:52	SMPL
K5390-02	MH0GS1	4540	1.0859	-	ISM02.4	PPB	10/30/2019 13:54	SMPL
K5390-03	MH0GS3	940	0.2176	-	ISM02.4	PPB	10/30/2019 13:57	SMPL
K5390-04	MH0GS5	493	0.1098	-	ISM02.4	PPB	10/30/2019 13:59	SMPL
K5390-05	MH0GS6	59	0.0051	-	ISM02.4	PPB	10/30/2019 14:01	SMPL
K5390-06	MH0GS7	31	-0.0016	-	ISM02.4	PPB	10/30/2019 14:03	SMPL
K5390-07	MH0GS9	285	0.0596	-	ISM02.4	PPB	10/30/2019 14:06	SMPL
K5390-08	MH0GT0	474	0.1052	-	ISM02.4	PPB	10/30/2019 14:08	SMPL
K5390-09	MH0GT1	117	0.0191	-	ISM02.4	PPB	10/30/2019 14:10	SMPL
K5390-10	MH0GT5	575	0.1296	-	ISM02.4	PPB	10/30/2019 14:13	SMPL

LB106018
INSTRUMENT ID : CV1

K5390-11	MH0GT6	98	0.0145 -	ISM02.4	PPB	10/30/2019 14:15 SMPL
K5390-12	MH0GT9	764	0.1752 -	ISM02.4	PPB	10/30/2019 14:17 SMPL
K5390-13	MH0GU1	109	0.0172 -	ISM02.4	PPB	10/30/2019 14:19 SMPL
K5390-14	MH0GU5	162	0.03 -	ISM02.4	PPB	10/30/2019 14:22 SMPL
K5390-15	MH0GU7	23	-0.0035 -	ISM02.4	PPB	10/30/2019 14:24 SMPL
K5390-16	MH0GU9	47	0.0022 -	ISM02.4	PPB	10/30/2019 14:26 SMPL
CCV056	CCV056	20324	4.893 -	ISM02.4	PPB	10/30/2019 14:29 SMPL
CCB056	CCB056	0	-0.0091 -	ISM02.4	PPB	10/30/2019 14:31 SMPL
K5390-17	MH0GV1	34	-0.0009 -	ISM02.4	PPB	10/30/2019 14:33 SMPL
K5390-18	MH0GV4	45	0.0018 -	ISM02.4	PPB	10/30/2019 14:35 SMPL
K5390-19	MH0GV5	1558	0.3667 -	ISM02.4	PPB	10/30/2019 14:38 SMPL
K5390-20	MH0GV8	110	0.0174 -	ISM02.4	PPB	10/30/2019 14:40 SMPL
K5390-21	MH0GV8D	101	0.0153 -	ISM02.4	PPB	10/30/2019 14:42 SMPL
K5390-22	MH0GV8S	4594	1.099 -	ISM02.4	PPB	10/30/2019 14:47 SMPL
PB124418E	PBW418	40	0.0006 -	ISM02.4	PPB	10/30/2019 14:49 SMPL
K5391-01	MH0GW0	243	0.0495 -	ISM02.4	PPB	10/30/2019 14:52 SMPL
K5391-02	MH0GW1	43	0.0013 -	ISM02.4	PPB	10/30/2019 14:55 SMPL
K5391-03	MH0GW2	15	-0.0055 -	ISM02.4	PPB	10/30/2019 14:57 SMPL
K5391-04	MH0GW3	361	0.078 -	ISM02.4	PPB	10/30/2019 14:59 SMPL
K5391-05	MH0GW4	32	-0.0014 -	ISM02.4	PPB	10/30/2019 15:02 SMPL
K5391-06	MH0GW8	132	0.0227 -	ISM02.4	PPB	10/30/2019 15:04 SMPL
K5391-07	MH0GX2	759	0.174 -	ISM02.4	PPB	10/30/2019 15:06 SMPL
K5391-08	MH0GX3	51	0.0032 -	ISM02.4	PPB	10/30/2019 15:09 SMPL
K5391-09	MH0GX5	685	0.1561 -	ISM02.4	PPB	10/30/2019 15:11 SMPL
K5391-10	MH0GX6	332	0.071 -	ISM02.4	PPB	10/30/2019 15:13 SMPL
K5391-11	MH0GX7	61	0.0056 -	ISM02.4	PPB	10/30/2019 15:15 SMPL
K5391-12	MH0GY1	118	0.0194 -	ISM02.4	PPB	10/30/2019 15:18 SMPL
K5391-13	MH0GY2	91	0.0129 -	ISM02.4	PPB	10/30/2019 15:20 SMPL
CCV057	CCV057	21393	5.1508 -	ISM02.4	PPB	10/30/2019 15:22 SMPL
CCB057	CCB057	8	-0.0072 -	ISM02.4	PPB	10/30/2019 15:25 SMPL
K5391-14	MH0GY3	92	0.0131 -	ISM02.4	PPB	10/30/2019 15:27 SMPL
K5391-15	MH0GY6	34	-0.0009 -	ISM02.4	PPB	10/30/2019 15:29 SMPL
K5391-16	MH0GZ5	185	0.0355 -	ISM02.4	PPB	10/30/2019 15:31 SMPL
K5391-17	MH0GZ5D	182	0.0348 -	ISM02.4	PPB	10/30/2019 15:34 SMPL
K5391-18	MH0GZ5S	4625	1.1064 -	ISM02.4	PPB	10/30/2019 15:36 SMPL
PB124416T	LEB009	30	-0.0019 -	ISM02.4	PPB	10/30/2019 15:38 SMPL
PB124417T	LEB009	31	-0.0016 -	ISM02.4	PPB	10/30/2019 15:41 SMPL
PB124418T	LEB009	56	0.0044 -	ISM02.4	PPB	10/30/2019 15:43 SMPL
CCV58 CCV58 CCV58	CCV58 CCV58 CCV58	21024	5.0618 -	ISM02.4	PPB	10/30/2019 15:45 SMPL
CCB58 CCB58 CCB58	CCB58 CCB58 CCB58	48	0.0025 -	ISM02.4	PPB	10/30/2019 15:47 SMPL