



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB106026

Slope : 98.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	10/30/2019	16:00
2	CAL2	1	Water	NA	NA	20.1	7.01	10/30/2019	16:01
3	CAL3	1	Water	NA	NA	20.1	10.03	10/30/2019	16:02
4	ICV	1	Water	NA	NA	20.1	7.00	10/30/2019	16:05
5	CCV1	1	Water	NA	NA	20.2	2.02	10/30/2019	16:06
6	K5612-01	1	Water	NA	NA	19.7	6.26	10/30/2019	16:37
7	K5612-01DUP	1	Water	NA	NA	19.8	6.27	10/30/2019	16:38
8	CCV2	1	Water	NA	NA	20.1	12.02	10/30/2019	16:40

2B106026

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K5615 W WorkList ID : 132782 Department : Wet-Chemistry Date : 10-30-2019 15:35:40

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/30/2019	Water	K5612-01 - <i>PH</i>	pH	Cool 4 deg C	PSEG01	N51	DRUM-COMP	10/30/2019	9040C

Date/Time 10.30.19 15.40
Received by: Jo
Relinquished by: cd

Date/Time 10.30.19 16.45
Received by: cd
Relinquished by: Jo