


7470A

LB106033

INSTRUMENT ID : CV1

Linear



A= 0.0000e+000

B= 2.4224e-004

C= 1.3195e-003

Rho= 0.9999546

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.019	0.019	73	0.000	73				
0.2	0.200	0.177	-0.023	726	0.0 %	726				
2.5	2.500	2.465	-0.035	10170	0.0 %	10170				
5.0	5.000	5.070	0.070	20922	0.0 %	20922				
7.5	7.500	7.483	-0.017	30883	0.0 %	30883				
10.0	10.000	9.987	-0.013	41221	0.0 %	41221				

LB106033 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 31 Oct 2019 09:22:28

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50-0	73	-	0.00007470A	PPB	31 Oct 2019 09:44:06	S	Std
0.2 - 1	50-2	726	-	0.20007470A	PPB	31 Oct 2019 09:46:15	S	Std
2.5 - 1	52-5	10170	-	2.50007470A	PPB	31 Oct 2019 09:48:24	S	Std
5.0 - 1	55-0	20922	-	5.00007470A	PPB	31 Oct 2019 09:50:33	S	Std
7.5 - 1	57-5	30883	-	7.50007470A	PPB	31 Oct 2019 09:52:41	S	Std
10.0 - 1	510-0	41221	-	10.00007470A	PPB	31 Oct 2019 09:57:16	S	Std
ICV82 - 1	ICV82	16536	4.0071	-7470A	PPB	31 Oct 2019 09:59:52	U	SMPL
ICB82 - 1	ICB82	56	0.0149	-7470A	PPB	31 Oct 2019 10:01:58	U	SMPL
CCV59 - 1	CCV59	21241	5.1468	-7470A	PPB	31 Oct 2019 10:04:05	U	SMPL
CCB59 - 1	CCB59	55	0.0146	-7470A	PPB	31 Oct 2019 10:06:12	U	SMPL
CRA - 1	CRA	742	0.1811	-7470A	PPB	31 Oct 2019 10:08:19	U	SMPL
HighStd - 1	HighStd	41529	10.0614	-7470A	PPB	31 Oct 2019 10:10:25	U	SMPL
ChkStd - 1	ChkStd	28533	6.9132	-7470A	PPB	31 Oct 2019 10:12:32	U	SMPL
PB124457BL - 1	PBW	103	0.0263	-7470A	PPB	31 Oct 2019 10:14:40	U	SMPL
PB124457BS - 1	LCSW	15272	3.7009	-7470A	PPB	31 Oct 2019 10:16:48	U	SMPL
K5149-12 - 1	EO-01-102819	71	0.0185	-7470A	PPB	31 Oct 2019 10:18:55	U	SMPL
K5149-12DUP - 1	EO-01-102819DUP	94	0.0241	-7470A	PPB	31 Oct 2019 10:21:03	U	SMPL
K5149-12MS - 1	EO-01-102819MS	19338	4.6858	-7470A	PPB	31 Oct 2019 10:23:11	U	SMPL
K5149-12MSD - 1	EO-01-102819MSD	18350	4.4465	-7470A	PPB	31 Oct 2019 10:25:20	U	SMPL
K5576-01 - 1	CSA-2-1-3	38	0.0105	-7470A	PPB	31 Oct 2019 10:27:27	U	SMPL
CCV60 - 1	CCV60	21136	5.1214	-7470A	PPB	31 Oct 2019 10:29:34	U	SMPL
CCB60 - 1	CCB60	163	0.0408	-7470A	PPB	31 Oct 2019 10:31:41	U	SMPL
K5576-03 - 1	CSA-2-1-4	34	0.0096	-7470A	PPB	31 Oct 2019 10:33:47	U	SMPL
K5576-05 - 1	CSA-2-1-5	47	0.0127	-7470A	PPB	31 Oct 2019 10:35:55	U	SMPL
K5576-07 - 1	CSA-2-1-6	52	0.0139	-7470A	PPB	31 Oct 2019 10:38:02	U	SMPL
K5576-09 - 1	CSA-2-1-7	60	0.0159	-7470A	PPB	31 Oct 2019 10:40:09	U	SMPL
K5576-11 - 1	CSA-2-1-8	59	0.0156	-7470A	PPB	31 Oct 2019 10:42:18	U	SMPL
K5587-03 - 1	CSA-2-1-9	47	0.0127	-7470A	PPB	31 Oct 2019 10:44:26	U	SMPL
K5587-06 - 1	CSA-2-1-10	37	0.0103	-7470A	PPB	31 Oct 2019 10:46:35	U	SMPL
K5587-09 - 1	CSA-2-1-11	47	0.0127	-7470A	PPB	31 Oct 2019 10:48:43	U	SMPL
K5587-12 - 1	CSA-2-1-12	32	0.0091	-7470A	PPB	31 Oct 2019 10:50:54	U	SMPL
PB124458BL - 1	PBW	61	0.0161	-7470A	PPB	31 Oct 2019 10:53:02	U	SMPL
CCV61 - 1	CCV61	20868	5.0565	-7470A	PPB	31 Oct 2019 10:55:10	U	SMPL
CCB61 - 1	CCB61	163	0.0408	-7470A	PPB	31 Oct 2019 10:57:17	U	SMPL
PB124458BS - 1	LCSW	15208	3.6854	-7470A	PPB	31 Oct 2019 10:59:26	U	SMPL
K5148-17 - 1	OR-03-102919	50	0.0134	-7470A	PPB	31 Oct 2019 11:01:33	U	SMPL
K5148-17DUP - 1	OR-03-102919DUP	-24	-0.0045	-7470A	PPB	31 Oct 2019 11:03:40	U	SMPL
K5148-17MS - 1	OR-03-102919MS	17090	4.1413	-7470A	PPB	31 Oct 2019 11:05:45	U	SMPL
K5148-17MSD - 1	OR-03-102919MSD	16921	4.1003	-7470A	PPB	31 Oct 2019 11:07:53	U	SMPL
K5599-02 - 1	Z3-20A	213	0.0529	-7470A	PPB	31 Oct 2019 11:10:00	U	SMPL
K5599-04 - 1	Z3-20B	62	0.0163	-7470A	PPB	31 Oct 2019 11:12:08	U	SMPL
K5599-06 - 1	Z3-21A	200	0.0498	-7470A	PPB	31 Oct 2019 11:14:17	U	SMPL
K5599-08 - 1	Z3-21B	189	0.0471	-7470A	PPB	31 Oct 2019 11:16:26	U	SMPL
K5599-10 - 1	Z3-22A	219	0.0544	-7470A	PPB	31 Oct 2019 11:18:32	U	SMPL
CCV62 - 1	CCV62	22121	5.3600	-7470A	PPB	31 Oct 2019 11:20:44	U	SMPL
CCB62 - 1	CCB62	117	0.0297	-7470A	PPB	31 Oct 2019 11:22:55	U	SMPL
K5599-12 - 1	Z3-22B	200	0.0498	-7470A	PPB	31 Oct 2019 11:25:04	U	SMPL
K5599-14 - 1	Z3-23A	196	0.0488	-7470A	PPB	31 Oct 2019 11:27:12	U	SMPL
K5599-16 - 1	Z3-23B	187	0.0466	-7470A	PPB	31 Oct 2019 11:29:19	U	SMPL
PB124459BL - 1	PBW	38	0.0105	-7470A	PPB	31 Oct 2019 11:31:27	U	SMPL
PB124459BS - 1	LCSW	15597	3.7796	-7470A	PPB	31 Oct 2019 11:33:34	U	SMPL
K5602-06 - 1	20190937-FIELD-BLANK	62	0.0163	-7470A	PPB	31 Oct 2019 11:35:42	U	SMPL
K5603-04 - 1	20190895-FIELD-BLANK	101	0.0258	-7470A	PPB	31 Oct 2019 11:37:51	U	SMPL
K5612-01 - 1	DRUM-COMP	43	0.0117	-7470A	PPB	31 Oct 2019 11:39:59	U	SMPL
K5612-01DUP - 1	DRUM-COMPDUP	60	0.0159	-7470A	PPB	31 Oct 2019 11:42:08	U	SMPL
K5612-01MS - 1	DRUM-COMPMS	16262	3.9407	-7470A	PPB	31 Oct 2019 11:44:15	U	SMPL
CCV63 - 1	CCV63	20970	5.0812	-7470A	PPB	31 Oct 2019 11:46:22	U	SMPL
CCB63 - 1	CCB63	7	0.0030	-7470A	PPB	31 Oct 2019 11:48:30	U	SMPL
K5612-01MSD - 1	DRUM-COMPMSD	16025	3.8833	-7470A	PPB	31 Oct 2019 11:50:37	U	SMPL
PB124457TB - 1	PB124370TB	153	0.0384	-7470A	PPB	31 Oct 2019 11:52:45	U	SMPL
PB124458TB - 1	PB124370TB	180	0.0449	-7470A	PPB	31 Oct 2019 11:54:54	U	SMPL
K5149-12LX5 - 1		23	0.0069	-7470A	PPB	31 Oct 2019 11:57:01	U	SMPL
K5149-12A - 1		18856	4.5691	-7470A	PPB	31 Oct 2019 11:59:09	U	SMPL
K5148-17LX5 - 1		2	0.0018	-7470A	PPB	31 Oct 2019 12:01:17	U	SMPL
K5148-17A - 1		17040	4.1291	-7470A	PPB	31 Oct 2019 12:03:26	U	SMPL
K5612-01LX5 - 1		-14	-0.0021	-7470A	PPB	31 Oct 2019 12:05:35	U	SMPL
K5612-01A - 1		16148	3.9131	-7470A	PPB	31 Oct 2019 12:07:43	U	SMPL
CCV64 - 1	CCV64	21261	5.1517	-7470A	PPB	31 Oct 2019 12:09:50	U	SMPL
CCB64 - 1	CCB64	6	0.0028	-7470A	PPB	31 Oct 2019 12:11:58	U	SMPL