



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB106037

Slope : 97.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	10/31/2019	13:15
2	CAL2	1	Water	NA	NA	20.1	7.00	10/31/2019	13:16
3	CAL3	1	Water	NA	NA	20.1	10.02	10/31/2019	13:18
4	ICV	1	Water	NA	NA	20.1	6.99	10/31/2019	13:20
5	CCV1	1	Water	NA	NA	20.1	2.02	10/31/2019	13:22
6	K5623-03	1	Water	NA	NA	20.2	7.00	10/31/2019	13:25
7	K5623-03DUP	1	Water	NA	NA	20.1	7.01	10/31/2019	13:26
8	CCV2	1	Water	NA	NA	20.1	12.02	10/31/2019	13:29

2B106038

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K5623 WorkList ID : 132823 Department : Wet-Chemistry Date : 10-31-2019 13:08:29

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/31/2019	Water	K5623-03 -18	Corrosivity	Cool 4 deg C	PSEG01	N52	PAT-TOTE	10/31/2019	9040C

Date/Time 10-31-19 13:10
Received by: Jd
Relinquished by: CA

Date/Time 10-31-19 14:00
Received by: CD
Relinquished by: Jd