

LB106038


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

56369



A= 0.0000e+000

B= 1.7672e-004

C= -1.5775e-002

Rho= 0.9999548

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.002	0.002	98	0.000	98				
0.2	0.200	0.171	-0.029	1059	0.0 %	1059				
2.5	2.500	2.502	0.002	14245	0.0 %	14245				
5.0	5.000	5.025	0.025	28521	0.0 %	28521				
7.5	7.500	7.555	0.055	42840	0.0 %	42840				
10.0	10.000	9.946	-0.054	56369	0.0 %	56369				

LB106038 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 31 Oct 2019 13:09:40				
Sample ID	Extended ID	µ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50.0	98	-	0.00007471B	PPB	31 Oct 2019 13:14:32	S	Std
0.2 - 1	50.2	1059	-	0.20007471B	PPB	31 Oct 2019 13:16:51	S	Std
2.5 - 1	52.5	14245	-	2.50007471B	PPB	31 Oct 2019 13:19:09	S	Std
5.0 - 1	55.0	28521	-	5.00007471B	PPB	31 Oct 2019 13:21:27	S	Std
7.5 - 1	57.5	42840	-	7.50007471B	PPB	31 Oct 2019 13:28:52	S	Std
10.0 - 1	57.5	56369	-	10.00007471B	PPB	31 Oct 2019 13:31:10	S	Std
ICV83 - 1	ICV83	24371	4.2911	-7471B	PPB	31 Oct 2019 13:33:46	U	SMPL
ICB83 - 1	ICB83	39	-0.0089	-7471B	PPB	31 Oct 2019 13:36:02	U	SMPL
CCV65 - 1	CCV65	28452	5.0123	-7471B	PPB	31 Oct 2019 13:38:18	U	SMPL
CCB65 - 1	CCB65	3	-0.0152	-7471B	PPB	31 Oct 2019 13:40:34	U	SMPL
CRA - 1	CRA	1064	0.1723	-7471B	PPB	31 Oct 2019 13:42:50	U	SMPL
HighStd - 1	HighStd	56914	10.0422	-7471B	PPB	31 Oct 2019 13:45:06	U	SMPL
ChkStd - 1	ChkStd	39753	7.0095	-7471B	PPB	31 Oct 2019 13:47:25	U	SMPL
PB124477BL - 1	PBS	82	-0.0013	-7471B	PPB	31 Oct 2019 13:49:43	U	SMPL
PB124477BS - 1	LCSS	21534	3.7898	-7471B	PPB	31 Oct 2019 13:52:02	U	SMPL
K5138-01 - 1	AU-08-103019	2315	0.3933	-7471B	PPB	31 Oct 2019 13:54:18	U	SMPL
K5138-01DUP - 1	AU-08-103019DUP	2371	0.4032	-7471B	PPB	31 Oct 2019 13:56:37	U	SMPL
K5138-01MS - 1	AU-08-103019MS	23987	4.2233	-7471B	PPB	31 Oct 2019 13:58:57	U	SMPL
K5138-01MSD - 1	AU-08-103019MSD	23518	4.1404	-7471B	PPB	31 Oct 2019 14:01:16	U	SMPL
K5143-05 - 1	NB-08-103019	2085	0.3527	-7471B	PPB	31 Oct 2019 14:03:33	U	SMPL
CCV66 - 1	CCV66	28555	5.0305	-7471B	PPB	31 Oct 2019 14:05:49	U	SMPL
CCB66 - 1	CCB66	34	-0.0098	-7471B	PPB	31 Oct 2019 14:08:05	U	SMPL
K5146-15 - 1	CL-01-103019	2071	0.3502	-7471B	PPB	31 Oct 2019 14:10:22	U	SMPL
K5147-08 - 1	OK-01-103019	1079	0.1749	-7471B	PPB	31 Oct 2019 14:12:38	U	SMPL
K5602-01 - 1	20190938-COMPOSITE-A	41968	7.4009	-7471B	PPB	31 Oct 2019 14:14:54	U	SMPL
K5602-02 - 1	20190939-COMPOSITE-B	29744	5.2407	-7471B	PPB	31 Oct 2019 14:17:11	U	SMPL
K5602-03 - 1	20190940-FIELD-DUP-COMPOSITE	35621	6.2793	-7471B	PPB	31 Oct 2019 14:19:29	U	SMPL
K5602-04 - 1	20190941-AMENDED-COMPOSITE-A	37535	6.6175	-7471B	PPB	31 Oct 2019 14:21:45	U	SMPL
K5602-05 - 1	20190942-AMENDED-COMPOSITE-B	25942	4.5688	-7471B	PPB	31 Oct 2019 14:24:03	U	SMPL
K5603-01 - 1	20190948-COMPOSITE-KM-PA	51067	9.0089	-7471B	PPB	31 Oct 2019 14:26:21	U	SMPL
K5603-02 - 1	20190949-AMND-COMPOSITE-KM-PA	40322	7.1100	-7471B	PPB	31 Oct 2019 14:28:40	U	SMPL
K5612-04 - 1	GL-STOCKPILE	1183	0.1933	-7471B	PPB	31 Oct 2019 14:30:58	U	SMPL
CCV67 - 1	CCV67	28406	5.0042	-7471B	PPB	31 Oct 2019 14:33:15	U	SMPL
CCB67 - 1	CCB67	31	-0.0103	-7471B	PPB	31 Oct 2019 14:35:31	U	SMPL
K5614-01 - 1	TREATMENT-SYSTEM-MEDIA	4640	0.8042	-7471B	PPB	31 Oct 2019 14:37:47	U	SMPL
K5615-01 - 1	RBR-29	1284	0.2111	-7471B	PPB	31 Oct 2019 14:41:34	U	SMPL
K5615-03 - 1	RBR-52	2381	0.4050	-7471B	PPB	31 Oct 2019 14:43:50	U	SMPL
K5615-05 - 1	3465	1519	0.2527	-7471B	PPB	31 Oct 2019 14:46:06	U	SMPL
K5620-03 - 1	TP-25-S	2752	0.4706	-7471B	PPB	31 Oct 2019 14:48:23	U	SMPL
K5620-07 - 1	TP-26A-S	1225	0.2007	-7471B	PPB	31 Oct 2019 14:50:40	U	SMPL
K5620-11 - 1	TP-26A-D	828	0.1306	-7471B	PPB	31 Oct 2019 14:52:57	U	SMPL
K5138-01LX5 - 1		474	0.0680	-7471B	PPB	31 Oct 2019 14:55:15	U	SMPL
K5138-01A - 1		23630	4.1602	-7471B	PPB	31 Oct 2019 14:57:33	U	SMPL
CCV68 - 1	CCV68	27672	4.8745	-7471B	PPB	31 Oct 2019 14:59:49	U	SMPL
CCB68 - 1	CCB68	22	-0.0119	-7471B	PPB	31 Oct 2019 15:02:05	U	SMPL