



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: AMANDEEP

Run Number: LB106041

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	10/31/2019	12:50
2	K5612-02		1	9	100.00	>212.0	>212.0	10/31/2019	13:20
3	K5612-02DUP		1	9	100.00	>212.0	>212.0	10/31/2019	13:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 106041

WorkList Name : FP103119 WorkList ID : 132820 Department : Wet-Chemistry Date : 10-31-2019 12:23:34

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/06/2019	Water	K5612-02	Flash Point	Cool 4 deg C	PSEG01	N51	DRUM-COMP	10/30/2019	1010A

Date/Time 10/31/19 12:25
 Received by: KS
 Relinquished by: CP

Date/Time 10/31/19 13:50
 Received by: CP
 Relinquished by: KS