

LB106055


7470A

INSTRUMENT ID: CU1

Linear

 μ Abs.:

39255



A= 0.0000e+000

B= 2.5545e-004

C= 8.5702e-003

Rho= 0.9999581

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.024	0.024	61	0.000	61				
0.2	0.200	0.180	-0.020	673	0.0 %	673				
2.5	2.500	2.537	0.037	9896	0.0 %	9896				
5.0	5.000	4.951	-0.049	19346	0.0 %	19346				
7.5	7.500	7.472	-0.028	29217	0.0 %	29217				
10.0	10.000	10.036	0.036	39255	0.0 %	39255				

LB106055 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 01 Nov 2019 09:28:01					
Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	61	-	0.0000	7470A	PPB	01 Nov 2019 10:59:24	S	Std
0.2 - 1	50.2	673	-	0.2000	7470A	PPB	01 Nov 2019 11:01:32	S	Std
2.5 - 1	52.5	9896	-	2.5000	7470A	PPB	01 Nov 2019 11:03:42	S	Std
5.0 - 1	55.0	19346	-	5.0000	7470A	PPB	01 Nov 2019 11:05:50	S	Std
7.5 - 1	57.5	29217	-	7.5000	7470A	PPB	01 Nov 2019 11:07:59	S	Std
10.0 - 1	510.0	39255	-	10.0000	7470A	PPB	01 Nov 2019 11:10:10	S	Std
ICV84 - 1	ICV84	16298	4.1719	-	7470A	PPB	01 Nov 2019 11:12:44	U	SMPL
ICB84 - 1	ICB84	4	0.0096	-	7470A	PPB	01 Nov 2019 11:14:52	U	SMPL
CCV69 - 1	CCV69	19136	4.8969	-	7470A	PPB	01 Nov 2019 11:16:58	U	SMPL
CCB69 - 1	CCB69	10	0.0111	-	7470A	PPB	01 Nov 2019 11:19:05	U	SMPL
CRA - 1	CRA	661	0.1774	-	7470A	PPB	01 Nov 2019 11:21:11	U	SMPL
HighStd - 1	HighStd	39183	10.0179	-	7470A	PPB	01 Nov 2019 11:23:17	U	SMPL
ChkStd - 1	ChkStd	27056	6.9200	-	7470A	PPB	01 Nov 2019 11:25:26	U	SMPL
PB124491BL - 1	PBW	-28	0.0014	-	7470A	PPB	01 Nov 2019 11:27:33	U	SMPL
PB124491BS - 1	LCSW	15815	4.0485	-	7470A	PPB	01 Nov 2019 11:29:40	U	SMPL
K5602-07 - 1	20190941-AMENDED-COMPOSITE-A	1165	0.3062	-	7470A	PPB	01 Nov 2019 11:31:47	U	SMPL
K5602-07DUP - 1	20190941-AMENDED-COMPOSITE-ADUP	1196	0.3141	-	7470A	PPB	01 Nov 2019 11:33:55	U	SMPL
K5602-07MS - 1	20190941-AMENDED-COMPOSITE-AMS	18503	4.7352	-	7470A	PPB	01 Nov 2019 11:36:02	U	SMPL
K5602-07MSD - 1	20190941-AMENDED-COMPOSITE-AMSD	17548	4.4912	-	7470A	PPB	01 Nov 2019 11:38:11	U	SMPL
K5602-08 - 1	20190942-AMENDED-COMPOSITE-B	1237	0.3246	-	7470A	PPB	01 Nov 2019 11:40:17	U	SMPL
CCV70 - 1	CCV70	19301	4.9390	-	7470A	PPB	01 Nov 2019 11:42:24	U	SMPL
CCB70 - 1	CCB70	2	0.0091	-	7470A	PPB	01 Nov 2019 11:44:31	U	SMPL
K5603-03 - 1	20190949-AMND-COMPOSITE-KM-PA	2590	0.6702	-	7470A	PPB	01 Nov 2019 11:46:37	U	SMPL
PB124492BL - 1	PBW	6	0.0101	-	7470A	PPB	01 Nov 2019 11:48:45	U	SMPL
PB124492BS - 1	LCSW	15941	4.0807	-	7470A	PPB	01 Nov 2019 11:50:52	U	SMPL
K5138-06 - 1	AU-08-103019	-1	0.0083	-	7470A	PPB	01 Nov 2019 11:52:59	U	SMPL
K5138-06DUP - 1	AU-08-103019DUP	25	0.0150	-	7470A	PPB	01 Nov 2019 11:55:06	U	SMPL
K5138-06MS - 1	AU-08-103019MS	15034	3.8490	-	7470A	PPB	01 Nov 2019 11:57:14	U	SMPL
K5138-06MSD - 1	AU-08-103019MSD	15033	3.8488	-	7470A	PPB	01 Nov 2019 11:59:21	U	SMPL
K5143-09 - 1	NB-08-103019	15	0.0124	-	7470A	PPB	01 Nov 2019 12:01:30	U	SMPL
K5146-22 - 1	CL-01-103019	29	0.0160	-	7470A	PPB	01 Nov 2019 12:03:38	U	SMPL
K5147-11 - 1	OK-01-103019	33	0.0170	-	7470A	PPB	01 Nov 2019 12:05:45	U	SMPL
CCV71 - 1	CCV71	19246	4.9250	-	7470A	PPB	01 Nov 2019 12:07:51	U	SMPL
CCB71 - 1	CCB71	-2	0.0081	-	7470A	PPB	01 Nov 2019 12:09:58	U	SMPL
K5607-02 - 1	OILY-DEBRIS	25	0.0150	-	7470A	PPB	01 Nov 2019 12:12:05	U	SMPL
K5612-06 - 1	GL-STOCKPILE	35	0.0175	-	7470A	PPB	01 Nov 2019 12:14:11	U	SMPL
K5614-02 - 1	TREATMENT-SYSTEM-MEDIA	57	0.0231	-	7470A	PPB	01 Nov 2019 12:16:18	U	SMPL
K5615-02 - 1	RBR-29	55	0.0226	-	7470A	PPB	01 Nov 2019 12:18:25	U	SMPL
K5615-04 - 1	RBR-52	45	0.0201	-	7470A	PPB	01 Nov 2019 12:20:31	U	SMPL
K5615-06 - 1	3465	47	0.0206	-	7470A	PPB	01 Nov 2019 12:22:39	U	SMPL
K5617-01 - 1	LB-1-2-3-5-COMP	55	0.0226	-	7470A	PPB	01 Nov 2019 12:24:46	U	SMPL
K5620-04 - 1	TP-25-S	31	0.0165	-	7470A	PPB	01 Nov 2019 12:26:54	U	SMPL
K5620-08 - 1	TP-26A-S	32	0.0167	-	7470A	PPB	01 Nov 2019 12:29:02	U	SMPL
K5620-12 - 1	TP-26A-D	29	0.0160	-	7470A	PPB	01 Nov 2019 12:31:09	U	SMPL
CCV72 - 1	CCV72	19053	4.8757	-	7470A	PPB	01 Nov 2019 12:33:16	U	SMPL
CCB72 - 1	CCB72	27	0.0155	-	7470A	PPB	01 Nov 2019 12:35:22	U	SMPL
PB124493BL - 1	PBW	35	0.0175	-	7470A	PPB	01 Nov 2019 12:37:29	U	SMPL
PB124493BS - 1	LCSW	15941	4.0807	-	7470A	PPB	01 Nov 2019 12:39:36	U	SMPL
K5607-01 - 1	OILY-WATER	67	0.0257	-	7470A	PPB	01 Nov 2019 12:41:43	U	SMPL
K5612-02 - 1	DRUM-COMP	45	0.0201	-	7470A	PPB	01 Nov 2019 12:43:49	U	SMPL
K5612-02DUP - 1	DRUM-COMPDUP	41	0.0190	-	7470A	PPB	01 Nov 2019 12:45:56	U	SMPL
K5612-02MS - 1	DRUM-COMPMS	15754	4.0329	-	7470A	PPB	01 Nov 2019 12:48:04	U	SMPL
K5612-02MSD - 1	DRUM-COMPMSD	15576	3.9875	-	7470A	PPB	01 Nov 2019 12:50:11	U	SMPL
PB124491TB - 1	PB124419TB	54	0.0224	-	7470A	PPB	01 Nov 2019 12:52:20	U	SMPL
PB124492TB - 1	PB124435TB	32	0.0167	-	7470A	PPB	01 Nov 2019 12:54:30	U	SMPL
PB124493TB - 1	PB124438TB	55	0.0226	-	7470A	PPB	01 Nov 2019 12:56:37	U	SMPL
CCV73 - 1	CCV73	19163	4.9038	-	7470A	PPB	01 Nov 2019 12:58:44	U	SMPL
CCB73 - 1	CCB73	14	0.0121	-	7470A	PPB	01 Nov 2019 13:00:51	U	SMPL
K5602-07LX5 - 1		262	0.0755	-	7470A	PPB	01 Nov 2019 13:02:57	U	SMPL
K5602-07A - 1		18246	4.6695	-	7470A	PPB	01 Nov 2019 13:05:04	U	SMPL
K5138-06LX5 - 1		-11	0.0058	-	7470A	PPB	01 Nov 2019 13:07:10	U	SMPL
K5138-06A - 1		15341	3.9274	-	7470A	PPB	01 Nov 2019 13:09:17	U	SMPL
K5612-02LX5 - 1		-25	0.0022	-	7470A	PPB	01 Nov 2019 13:11:24	U	SMPL
K5612-02A - 1		15605	3.9949	-	7470A	PPB	01 Nov 2019 13:13:31	U	SMPL
CCV74 - 1	CCV74	19418	4.9689	-	7470A	PPB	01 Nov 2019 13:15:39	U	SMPL
CCB74 - 1	CCB74	17	0.0129	-	7470A	PPB	01 Nov 2019 13:17:47	U	SMPL