



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB106061

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	11/01/2019	12:15
2	CAL2	1	Water	NA	NA	20.1	7.00	11/01/2019	12:16
3	CAL3	1	Water	NA	NA	20.1	10.03	11/01/2019	12:17
4	ICV	1	Water	NA	NA	20.2	7.00	11/01/2019	12:19
5	CCV1	1	Water	NA	NA	20.2	2.01	11/01/2019	12:20
6	K5654-01	1	Water	NA	NA	18.7	4.60	11/01/2019	12:21
7	K5654-01DUP	1	Water	NA	NA	18.6	4.61	11/01/2019	12:22
8	CCV2	1	Water	NA	NA	20.2	12.02	11/01/2019	12:25

23106069

WORKLIST(Hardcopy Internal Chain)

WorkList Name :	PH K5652	WorkList ID :	132848	Department :	Wet-Chemistry	Date :	11-01-2019 12:05:08		
Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/30/2019	Water	K5654-01 - A	pH	Cool 4 deg C	STER14	H31	MINI-CELL	10/30/2019	SM 4500-H B

Date/Time 11-01-19 12:10
Received by: Jg
Relinquished by: Jg

Date/Time 11-01-19 13:00
Received by: Jg
Relinquished by: Jg