

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : *ET*

Instrument ID : Konelab

11/1/2019 13:46


Test Total CN

Accepted 11/1/2019 10:48

Factor 1186
Bias 0.002

Coeff. of det. 0.999603

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.002	-0.7896	0.0000	
2	5.0PPBCN	0.006	4.0150	5.0000	
3	10PPBCN	0.010	9.6856	10.0000	
4	50PPBCN	0.043	48.0850	50.0000	
5	100PPBCN	0.086	99.8039	100.0000	
6	250PPBCN	0.220	257.9066	250.0000	
7	500PPBCN	0.421	496.2936	500.0000	

LB106064

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092

Reviewed by : g

Instrument ID : Konelab

11/1/2019 14:13

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	98.050	0.0	0.085	
ICB1	-0.096	0.0	0.002	
CCV1	250.896	0.0	0.214	
CCB1	-0.226	0.0	0.002	
PB124462BL	-0.037	0.0	0.002	
PB124462BS	101.563	0.0	0.088	
K5609-05	2738.161	0.0	2.311	Test limit high
K5609-05DLX10	276.531	0.0	0.235	
CCV2	263.618	0.0	0.225	
CCB2	-0.130	0.0	0.002	

N	10
Mean	372.833
SD	839.0685
CV%	225.05

11/1/19
CT

Aquakem v. 7.2AQ1

Results from time period:

Fri Nov 01 14:04:12 2019

Fri Nov 01 14:08:34 2019

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	-0.7896	µg/l	11/1/2019 10:46:47	
5.OPPBCN	A	Total CN	P	4.015	µg/l	11/1/2019 10:46:48	
10PPBCN	A	Total CN	P	9.6856	µg/l	11/1/2019 10:46:49	
50PPBCN	A	Total CN	P	48.085	µg/l	11/1/2019 10:46:50	
100PPBCN	A	Total CN	P	99.8039	µg/l	11/1/2019 10:46:51	
250PPBCN	A	Total CN	P	257.9066	µg/l	11/1/2019 10:46:52	
500PPBCN	A	Total CN	P	496.2936	µg/l	11/1/2019 10:46:53	
ICV1	S	Total CN	P	98.0501	µg/l	11/1/2019 14:04:13	
ICB1	S	Total CN	P	-0.0957	µg/l	11/1/2019 14:04:14	
CCV1	S	Total CN	P	250.896	µg/l	11/1/2019 14:04:17	
CCB1	S	Total CN	P	-0.2262	µg/l	11/1/2019 14:04:19	
PB124462BL	S	Total CN	P	-0.0367	µg/l	11/1/2019 14:04:20	
PB124462BS	S	Total CN	P	101.5632	µg/l	11/1/2019 14:04:21	
K5609-05	S	Total CN	P	2738.161	µg/l	11/1/2019 14:04:22	
K5609-05DLX10	S	Total CN	P	276.5307	µg/l	11/1/2019 14:08:31	
CCV2	S	Total CN	P	263.6179	µg/l	11/1/2019 14:08:33	
CCB2	S	Total CN	P	-0.1301	µg/l	11/1/2019 14:08:34	