



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB106066

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	11/01/2019	14:50
2	K5623-03		1	9	100.00	>212.0	>212.0	11/01/2019	15:20
3	K5623-03DUP		1	9	100.00	>212.0	>212.0	11/01/2019	15:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB106066

WorkList Name : FP11119 WorkList ID : 132854 Department : Wet-Chemistry Date : 11-01-2019 14:20:33

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/06/2019	Water	K5623-03	Flash Point	Cool 4 deg C	PSEG01	N52	PAT-TOTE	10/31/2019	1010A

Date/Time 11/01/19 14:25
 Received by: KS
 Relinquished by: CP

Date/Time 11/01/19 15:50
 Received by: CP
 Relinquished by: KS