



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB106067

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	11/01/2019	16:15
2	CAL2	1	Water	NA	NA	20.1	7.00	11/01/2019	16:16
3	CAL3	1	Water	NA	NA	20.1	10.03	11/01/2019	16:17
4	ICV	1	Water	NA	NA	20.1	6.99	11/01/2019	16:19
5	CCV1	1	Water	NA	NA	20.1	2.02	11/01/2019	16:20
6	K5609-02	1	Solid	20.02	20	22.2	7.44	11/01/2019	16:21
7	K5658-04	1	Solid	20.02	20	22.3	7.12	11/01/2019	16:24
8	K5658-08	1	Solid	20.03	20	22.1	8.01	11/01/2019	16:25
9	K5676-05	1	Solid	20.04	20	20.9	7.95	11/01/2019	16:30
10	K5676-05DUP	1	Solid	20.03	20	20.8	7.96	11/01/2019	16:31
11	CCV2	1	Water	NA	NA	20.1	12.02	11/01/2019	16:32

VB 106067

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K5658 WorkList ID : 132851 Department : Wet-Chemistry Date : 11-01-2019 13:37:38

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/01/2019	Solid	K5658-04	Corrosivity	Cool 4 deg C	PSEG01	N52	TP-30-S	11/01/2019	9045D
11/01/2019	Solid	K5658-08	Corrosivity	Cool 4 deg C	PSEG01	N52	TP-30-D	11/01/2019	9045D
11/01/2019	Solid	K5676-05	Corrosivity	Cool 4 deg C	PSEG05	N52	SU-02-110119	11/01/2019	9045D

Date/Time 11-01-19 16:10
Received by: Jp
Relinquished by: ea

Date/Time 11-01-19 17:30
Received by: SD
Relinquished by: Jp