



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB106074

Slope : 99.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	11/04/2019	10:40
2	CAL2	1	Water	NA	NA	20.1	7.00	11/04/2019	10:41
3	CAL3	1	Water	NA	NA	20.1	10.03	11/04/2019	10:42
4	ICV	1	Water	NA	NA	20.1	7.01	11/04/2019	10:45
5	CCV1	1	Water	NA	NA	20.2	2.01	11/04/2019	10:46
6	K5680-01	1	Water	NA	NA	19.5	5.01	11/04/2019	10:47
7	K5680-01DUP	1	Water	NA	NA	19.6	5.02	11/04/2019	10:48
8	CCV2	1	Water	NA	NA	20.2	12.03	11/04/2019	10:49

LB 106034

WORKLIST(Hardcopy Internal Chain)

WorkList Name :	PH K5680	WorkList ID :	132884	Department :	Wet-Chemistry	Date :	11-04-2019 10:34:45		
Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/10/2019	Water	K5680-01-A	pH	Cool 4 deg C	STER10	H41	CELL-A-POOL-WATER(DE-IC	10/10/2019	SM 4500-H B

Date/Time 11-04-19 10:35
Received by: JB
Relinquished by: SQ

Date/Time 11-04-19 11:30
Received by: CR
Relinquished by: JN