



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: AMANDEEP

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB106075

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP78936
COD calibration std. 100 ppm	WP78932
COD calibration std. 10 ppm	WP78931
COD calibration std. 150 ppm	WP78933
COD calibration std. 50 ppm	WP78934
COD calibration std. 0 ppm	WP78927
COD CCV std, 50ppm	WP78935
COD Digestion Vials Low Level 0-150Mg/L	W2558
COD CCV std, 50ppm	WP79563
COD ICV-LCS std, 50ppm	WP79562

Temp In (C): 150	Date In: 11/04/2019	Time In: 08:30
Temp Out (C): 151	Date Out: 11/04/2019	Time Out: 10:30

Intercept: -0.3904

Slope: 1.0063

Regression: 1

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	10/15/2019	10:45
2	CAL2	10	1	Water	10.000	10/15/2019	10:45
3	CAL3	50	1	Water	49.000	10/15/2019	10:46
4	CAL4	100	1	Water	100.000	10/15/2019	10:46
5	CAL5	150	1	Water	151.000	10/15/2019	10:47

Analysis Method: SM5220 DANALYST: AMANDEEPParameter: CODSUPERVISOR REVIEW BY: apatelRun Number: LB106075

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	50.000	50.075	10/15/2019	10:47
2	ICB				1	Water	0.000	0.388	10/15/2019	10:48
3	CCV1	50			1	Water	50.000	50.075	11/04/2019	11:30
4	CCB1				1	Water	0.000	0.388	11/04/2019	11:30
5	LB106075BL				1	Water	0.000	0.388	11/04/2019	11:31
6	LB106075BS	50			1	Water	53.000	53.056	11/04/2019	11:31
7	K5608-01				1	Water	0.000	0.388	11/04/2019	11:32
8	K5608-02				1	Water	0.000	0.388	11/04/2019	11:32
9	K5608-02DUP				1	Water	0.000	0.388	11/04/2019	11:33
10	K5608-02MS	50			1	Water	49.000	49.081	11/04/2019	11:34
11	K5608-02MSD	50			1	Water	50.000	50.075	11/04/2019	11:34
12	K5612-01				1	Water	26.000	26.225	11/04/2019	11:35
13	K5623-02				1	Water	12.000	12.313	11/04/2019	11:35
14	CCV2	50			1	Water	49.000	49.081	11/04/2019	11:36
15	CCB2				1	Water	0.000	0.388	11/04/2019	11:36