

LB106081


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

39856



A= 0.0000e+000

B= 2.5017e-004

C= -5.8949e-002

Rho= 0.9998138

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.041	-0.041	73	0.000	73				
0.2	0.200	0.165	-0.035	896	0.0 %	896				
2.5	2.500	2.504	0.004	10246	0.0 %	10246				
5.0	5.000	5.139	0.139	20778	0.0 %	20778				
7.5	7.500	7.520	0.020	30297	0.0 %	30297				
10.0	10.000	9.912	-0.088	39856	0.0 %	39856				

LB106081 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 04 Nov 2019 12:56:34

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	73	-	0.0000	7470A	PPB	04 Nov 2019 13:02:10	S	Std
0.2 - 1	50-2	896	-	0.2000	7470A	PPB	04 Nov 2019 13:04:19	S	Std
2.5 - 1	52-5	10246	-	2.5000	7470A	PPB	04 Nov 2019 13:06:27	S	Std
5.0 - 1	55-0	20778	-	5.0000	7470A	PPB	04 Nov 2019 13:08:36	S	Std
7.5 - 1	57-5	30297	-	7.5000	7470A	PPB	04 Nov 2019 13:10:43	S	Std
10.0 - 1	510-0	39856	-	10.0000	7470A	PPB	04 Nov 2019 13:12:50	S	Std
ICV85 - 1	ICV85	16409	4.0461	-	7470A	PPB	04 Nov 2019 13:17:56	U	SMPL
ICB85 - 1	ICB85	24	-0.0529	-	7470A	PPB	04 Nov 2019 13:20:02	U	SMPL
CCV75 - 1	CCV75	20995	5.1933	-	7470A	PPB	04 Nov 2019 13:22:08	U	SMPL
CCB75 - 1	CCB75	16	-0.0549	-	7470A	PPB	04 Nov 2019 13:24:14	U	SMPL
CRA - 1	CRA	911	0.1690	-	7470A	PPB	04 Nov 2019 13:26:20	U	SMPL
HighStd - 1	HighStd	39504	9.8237	-	7470A	PPB	04 Nov 2019 13:28:26	U	SMPL
ChkStd - 1	ChkStd	28458	7.0603	-	7470A	PPB	04 Nov 2019 13:30:32	U	SMPL
PB124524BL - 1	PBW	32	-0.0509	-	7470A	PPB	04 Nov 2019 13:32:39	U	SMPL
PB124524BS - 1	LCSW	16192	3.9918	-	7470A	PPB	04 Nov 2019 13:34:46	U	SMPL
K5623-03 - 1	PAT-TOTE	38	-0.0494	-	7470A	PPB	04 Nov 2019 13:36:54	U	SMPL
K5623-03DUP - 1	PAT-TOTEDUP	64	-0.0429	-	7470A	PPB	04 Nov 2019 13:39:01	U	SMPL
K5623-03MS - 1	PAT-TOTEMS	16539	4.0786	-	7470A	PPB	04 Nov 2019 13:41:09	U	SMPL
K5623-03MSD - 1	PAT-TOTEMSD	16559	4.0836	-	7470A	PPB	04 Nov 2019 13:43:17	U	SMPL
PB124525BL - 1	PBW	53	-0.0457	-	7470A	PPB	04 Nov 2019 13:45:23	U	SMPL
CCV76 - 1	CCV76	20651	5.1073	-	7470A	PPB	04 Nov 2019 13:47:30	U	SMPL
CCB76 - 1	CCB76	51	-0.0462	-	7470A	PPB	04 Nov 2019 13:49:36	U	SMPL
PB124525BS - 1	LCSW	15496	3.8177	-	7470A	PPB	04 Nov 2019 13:51:42	U	SMPL
K5657-02 - 1	TP	69	-0.0417	-	7470A	PPB	04 Nov 2019 13:53:49	U	SMPL
K5658-04 - 1	TP-30-S	83	-0.0382	-	7470A	PPB	04 Nov 2019 13:55:56	U	SMPL
K5658-08 - 1	TP-30-D	76	-0.0399	-	7470A	PPB	04 Nov 2019 13:58:03	U	SMPL
K5676-05 - 1	SU-02-110119	47	-0.0472	-	7470A	PPB	04 Nov 2019 14:00:09	U	SMPL
K5676-05DUP - 1	SU-02-110119DUP	87	-0.0372	-	7470A	PPB	04 Nov 2019 14:02:17	U	SMPL
K5676-05MS - 1	SU-02-110119MS	16469	4.0611	-	7470A	PPB	04 Nov 2019 14:04:24	U	SMPL
K5676-05MSD - 1	SU-02-110119MSD	15951	3.9315	-	7470A	PPB	04 Nov 2019 14:06:32	U	SMPL
PB124526BL - 1	PBW	69	-0.0417	-	7470A	PPB	04 Nov 2019 14:08:40	U	SMPL
PB124526BS - 1	LCSW	16067	3.9605	-	7470A	PPB	04 Nov 2019 14:10:47	U	SMPL
CCV77 - 1	CCV77	20500	5.0695	-	7470A	PPB	04 Nov 2019 14:12:54	U	SMPL
CCB77 - 1	CCB77	51	-0.0462	-	7470A	PPB	04 Nov 2019 14:15:01	U	SMPL
K5623-02 - 1	PAT-TOTE	3004	0.6926	-	7470A	PPB	04 Nov 2019 14:17:07	U	SMPL
K5623-02DUP - 1	PAT-TOTEDUP	2779	0.6363	-	7470A	PPB	04 Nov 2019 14:19:14	U	SMPL
K5623-02MS - 1	PAT-TOTEMS	19780	4.8894	-	7470A	PPB	04 Nov 2019 14:21:21	U	SMPL
K5623-02MSD - 1	PAT-TOTEMSD	20038	4.9539	-	7470A	PPB	04 Nov 2019 14:23:28	U	SMPL
PB124524TB - 1	PB124499TB	17	-0.0547	-	7470A	PPB	04 Nov 2019 14:25:36	U	SMPL
PB124525TB - 1	PB124500TB	62	-0.0434	-	7470A	PPB	04 Nov 2019 14:27:43	U	SMPL
K5623-03LX5 - 1		74	-0.0404	-	7470A	PPB	04 Nov 2019 14:29:51	U	SMPL
K5623-03A - 1		16527	4.0756	-	7470A	PPB	04 Nov 2019 14:31:59	U	SMPL
K5676-05LX5 - 1		21	-0.0537	-	7470A	PPB	04 Nov 2019 14:34:07	U	SMPL
K5676-05A - 1		17237	4.2532	-	7470A	PPB	04 Nov 2019 14:36:14	U	SMPL
CCV78 - 1	CCV78	21047	5.2063	-	7470A	PPB	04 Nov 2019 14:38:21	U	SMPL
CCB78 - 1	CCB78	3	-0.0582	-	7470A	PPB	04 Nov 2019 14:40:27	U	SMPL
K5623-02LX5 - 1		602	0.0917	-	7470A	PPB	04 Nov 2019 14:42:34	U	SMPL
K5623-02A - 1		20147	4.9812	-	7470A	PPB	04 Nov 2019 14:44:40	U	SMPL
CCV79 - 1	CCV79	20740	5.1295	-	7470A	PPB	04 Nov 2019 14:46:47	U	SMPL
CCB79 - 1	CCB79	31	-0.0512	-	7470A	PPB	04 Nov 2019 14:48:53	U	SMPL