



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB106089

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	11/04/2019	14:15
2	CAL2	1	Water	NA	NA	20.1	7.01	11/04/2019	14:16
3	CAL3	1	Water	NA	NA	20.1	10.04	11/04/2019	14:17
4	ICV	1	Water	NA	NA	20.2	7.00	11/04/2019	14:19
5	CCV1	1	Water	NA	NA	20.1	2.01	11/04/2019	14:20
6	K5683-03	1	Water	NA	NA	17.2	6.96	11/04/2019	14:21
7	K5683-03DUP	1	Water	NA	NA	17.1	6.97	11/04/2019	14:22
8	CCV2	1	Water	NA	NA	20.1	12.01	11/04/2019	14:25

1B 106089

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY K5683 WorkList ID : 132889 Department : Wet-Chemistry Date : 11-04-2019 14:08:32

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/04/2019	Water	K5683-03- ☒	Corrosivity	Cool 4 deg C	PSEG01	N52	OILY-WATER	11/04/2019	9040C

Date/Time 11-04-19 14:10
 Received by: JD
 Relinquished by: CD

Date/Time 11-04-19 15:00
 Received by: CD
 Relinquished by: JD