

LB106103


7471B

INSTRUMENT ID : CV1

Linear

 μ Abs.:

57804



A= 0.0000e+000

B= 1.7420e-004

C= -5.2999e-002

Rho= 0.9999641

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.003	-0.003	287	0.000	287				
0.2	0.200	0.179	-0.021	1330	0.0 %	1330				
2.5	2.500	2.527	0.027	14808	0.0 %	14808				
5.0	5.000	5.037	0.037	29219	0.0 %	29219				
7.5	7.500	7.445	-0.055	43041	0.0 %	43041				
10.0	10.000	10.016	0.016	57804	0.0 %	57804				

LB106103 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 05 Nov 2019 10:34:03

Sample ID	Extended ID	µ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	287	-	0.0000	7471B	PPB	05 Nov 2019 10:50:03	S	Std
0.2 - 1	50.2	1330	-	0.2000	7471B	PPB	05 Nov 2019 10:52:29	S	Std
2.5 - 1	52.5	14808	-	2.5000	7471B	PPB	05 Nov 2019 10:54:46	S	Std
5.0 - 1	55.0	29219	-	5.0000	7471B	PPB	05 Nov 2019 10:57:04	S	Std
7.5 - 1	57.5	43041	-	7.5000	7471B	PPB	05 Nov 2019 11:05:47	S	Std
10.0 - 1	516.0	57804	-	10.0000	7471B	PPB	05 Nov 2019 11:08:04	S	Std
ICV86 - 1	ICV86	23095	3.9701	-	7471B	PPB	05 Nov 2019 11:11:57	U	SMPL
ICB86 - 1	ICB86	232	-0.0126	-	7471B	PPB	05 Nov 2019 11:14:12	U	SMPL
CCV80 - 1	CCV80	29608	5.1046	-	7471B	PPB	05 Nov 2019 11:16:27	U	SMPL
CCB80 - 1	CCB80	226	-0.0136	-	7471B	PPB	05 Nov 2019 11:18:42	U	SMPL
CRA - 1	CRA	1369	0.1855	-	7471B	PPB	05 Nov 2019 11:20:58	U	SMPL
HighStd - 1	HighStd	58416	10.1229	-	7471B	PPB	05 Nov 2019 11:23:13	U	SMPL
ChkStd - 1	ChkStd	39953	6.9067	-	7471B	PPB	05 Nov 2019 11:25:28	U	SMPL
PB124547BL - 1	PB124547BL	222	-0.0143	-	7471B	PPB	05 Nov 2019 11:27:47	U	SMPL
PB124547BS - 1	LCSS	21628	3.7145	-	7471B	PPB	05 Nov 2019 11:30:04	U	SMPL
K5609-11 - 1	PT-MET-SOIL	700268	HIGH	-	7471B	PPB	05 Nov 2019 11:32:21	X	
PB124548BL - 1	PB124548BL	-571	-0.1525	-	7471B	PPB	05 Nov 2019 11:34:12	U	SMPL
PB124548BS - 1	PB124548BS	21589	3.7077	-	7471B	PPB	05 Nov 2019 11:36:33	U	SMPL
K5624-01 - 1	WASTE	706826	HIGH	-	7471B	PPB	05 Nov 2019 11:38:50	X	
K5624-01DUP - 1	WASTEDUP	847545	HIGH	-	7471B	PPB	05 Nov 2019 11:40:39	X	
CCV81 - 1	CCV81	28979	4.9951	-	7471B	PPB	05 Nov 2019 11:42:32	U	SMPL
CCB81 - 1	CCB81	-111	-0.0723	-	7471B	PPB	05 Nov 2019 11:44:59	U	SMPL
K5628-01 - 1	RBR200055-RBR200034-RT1868	4031	0.6492	-	7471B	PPB	05 Nov 2019 11:50:56	U	SMPL
K5657-01 - 1	TP	12806	2.1778	-	7471B	PPB	05 Nov 2019 11:53:19	U	SMPL
K5658-03 - 1	TP-30-S	1034	0.1271	-	7471B	PPB	05 Nov 2019 11:55:36	U	SMPL
K5658-07 - 1	TP-30-D	326	0.0038	-	7471B	PPB	05 Nov 2019 11:57:52	U	SMPL
K5662-01 - 1	S-1	1495	0.2074	-	7471B	PPB	05 Nov 2019 12:00:09	U	SMPL
K5676-01 - 1	SU-02-110119	1399	0.1907	-	7471B	PPB	05 Nov 2019 12:02:26	U	SMPL
K5624-01LX5 - 1		345310	60.0991	-	7471B	PPB	05 Nov 2019 12:04:43	U	SMPL
K5624-01A - 1		695554	HIGH	-	7471B	PPB	05 Nov 2019 12:06:59	X	
CCV82 - 1	CCV82	28919	4.9846	-	7471B	PPB	05 Nov 2019 12:08:53	U	SMPL
CCB82 - 1	CCB82	-107	-0.0716	-	7471B	PPB	05 Nov 2019 12:11:12	U	SMPL
K5624-01DLX50 - 1		32576	5.6217	-	7471B	PPB	05 Nov 2019 12:19:59	U	SMPL
K5624-01DUPDLX50 - 1		48887	8.4636	-	7471B	PPB	05 Nov 2019 12:22:17	U	SMPL
K5624-01MS - 1	WASTEMS	854460	HIGH	-	7471B	PPB	05 Nov 2019 12:24:35	X	
K5624-01MSD - 1	WASTEMSD	717851	HIGH	-	7471B	PPB	05 Nov 2019 12:26:26	X	
K5624-01ADLX50 - 1		48018	8.3116	-	7471B	PPB	05 Nov 2019 12:30:49	U	SMPL
CCV83 - 1	CCV83	28065	4.8358	-	7471B	PPB	05 Nov 2019 12:35:24	U	SMPL
CCB83 - 1	CCB83	-52	-0.0621	-	7471B	PPB	05 Nov 2019 12:37:42	U	SMPL
K5609-11DLX40 - 1		38979	6.7370	-	7471B	PPB	05 Nov 2019 12:44:05	U	SMPL
K5624-01DLX50 - 1		40136	6.9386	-	7471B	PPB	05 Nov 2019 12:46:28	U	SMPL
K5624-01DUPDLX50 - 1		39878	6.8936	-	7471B	PPB	05 Nov 2019 12:48:46	U	SMPL
K5624-01MSDLX50 - 1		47538	8.2280	-	7471B	PPB	05 Nov 2019 12:51:04	U	SMPL
K5624-01MSDDLX50 - 1		48105	8.3268	-	7471B	PPB	05 Nov 2019 12:53:22	U	SMPL
K5624-01DL2X250 - 1		7868	1.3176	-	7471B	PPB	05 Nov 2019 12:55:42	U	SMPL
CCV84 - 1	CCV84	28552	4.9207	-	7471B	PPB	05 Nov 2019 12:59:51	U	SMPL
CCB84 - 1	CCB84	-20	-0.0565	-	7471B	PPB	05 Nov 2019 13:02:15	U	SMPL

RS 11-05-19