



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB106108

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	3.99	11/05/2019	17:15
2	CAL2	1	Water	NA	NA	20.2	7.00	11/05/2019	17:16
3	CAL3	1	Water	NA	NA	20.2	10.03	11/05/2019	17:17
4	ICV	1	Water	NA	NA	20.2	7.01	11/05/2019	17:19
5	CCV1	1	Water	NA	NA	20.1	2.02	11/05/2019	17:20
6	K5671-05	1	Solid	20.03	20	21.4	8.60	11/05/2019	17:21
7	K5675-04	1	Solid	20.04	20	21.3	8.52	11/05/2019	17:22
8	K5675-04DUP	1	Solid	20.03	20	21.4	8.53	11/05/2019	17:24
9	CCV2	1	Water	NA	NA	20.1	12.02	11/05/2019	17:25

VB106108

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K5671 WorkList ID : 132924 Department : Wet-Chemistry Date : 11-05-2019 15:47:23

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/05/2019	Solid	K5671-05-9	Corrosivity	Cool 4 deg C	PSEG05	N52	OR-02-110519	11/05/2019	9045D
11/05/2019	Solid	K5675-04-9	Corrosivity	Cool 4 deg C	PSEG05	N52	TR-05-110519	11/05/2019	9045D

Date/Time 11-05-19 16:30
Received by: JH
Relinquished by: PC

Date/Time 11-05-19 18:10
Received by: PC
Relinquished by: JH