



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB106114

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	11/05/2019	15:45
2	CAL2	1	Water	NA	NA	20.2	7.01	11/05/2019	15:46
3	CAL3	1	Water	NA	NA	20.2	10.03	11/05/2019	15:47
4	ICV	1	Water	NA	NA	20.1	7.00	11/05/2019	15:48
5	CCV1	1	Water	NA	NA	20.1	2.01	11/05/2019	15:49
6	K5703-01	1	Water	NA	NA	16.1	7.14	11/05/2019	15:50
7	K5703-01DUP	1	Water	NA	NA	16.2	7.15	11/05/2019	15:51
8	CCV2	1	Water	NA	NA	20.1	12.01	11/05/2019	16:00

LB 106114

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K5703 WorkList ID : 132921 Department : Wet-Chemistry Date : 11-05-2019 15:26:26

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/05/2019	Water	K5703-01	pH	Cool 4 deg C	PSEG01	N52	OILY-WATER	11/05/2019	9040C

Date/Time 11-05-19 15:30
Received by: Jn
Relinquished by: [Signature]

Date/Time 11-05-19 16:30
Received by: [Signature]
Relinquished by: Jn