



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB106117

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	11/06/2019	11:10
2	CAL2	1	Water	NA	NA	20.1	7.00	11/06/2019	11:11
3	CAL3	1	Water	NA	NA	20.1	10.03	11/06/2019	11:12
4	ICV	1	Water	NA	NA	20.1	7.01	11/06/2019	11:15
5	CCV1	1	Water	NA	NA	20.1	2.02	11/06/2019	11:16
6	K5704-01	1	Water	NA	NA	6.8	6.67	11/06/2019	11:19
7	K5704-01DUP	1	Water	NA	NA	6.7	6.68	11/06/2019	11:20
8	K5704-04	1	Water	NA	NA	7.2	6.65	11/06/2019	11:23
9	CCV2	1	Water	NA	NA	20.1	12.02	11/06/2019	11:25

LB106117

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K5704 WorkList ID : 132947 Department : Wet-Chemistry Date : 11-06-2019 10:48:56

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/05/2019	Water	K5704-01 ✓	pH	Cool 4 deg C	APT01	H41	WGGM-19-WW-TSF-039	11/05/2019	SM 4500-H B
11/05/2019	Water	K5704-04 ✓	pH	Cool 4 deg C	APT01	H41	WGGM-19-WW-TSF-040	11/05/2019	SM 4500-H B

Date/Time 11-06-19 11:00
Received by: jo
Relinquished by: jo

Date/Time 11-06-19 11:45
Received by: aw
Relinquished by: jo