



Analytical Summary Report

Analysis Method: 7196A
Parameter: Hexavalent Chromium
Run Number: LB106118

ANALYST: ketankumar
SUPERVISOR REVIEW BY: apatel
pH Meter ID: WC pH Meter-1

Table with 2 columns: Reagent/Standard, Lot/Log #. Rows include various calibration standards and reagents like Hexachrome, Hexavalent Chromium ICV Std, and 5N sulfuric acid.

Intercept: -0.0002 Slope: 0.798 Regression: 0.999996

Table with 12 columns: Seq, Lab ID, True Value (mg/l), DF, Initial Vol (ml/gm), Final Vol (ml), pH HN03, pH H2SO4, Absorb. at 540 (Backgrnd, Color), Absorbance Difference, Anal Date, Anal Time. Contains 7 rows of calibration data.

Analysis Method: 7196A

ANALYST: ketankumar

Parameter: Hexavalent Chromiu

SUPERVISOR REVIEW BY: apatel

Run Number: LB106118

pH Meter ID: WC pH Meter-1

Seq	Lab ID	True Value (mg/l)	DF	Initial Vol (ml/gm)	Final Vol (ml)	pH HN03	pH H2SO4	Absorb. at 540		Absorbance Difference	Result	Anal Date	Anal Time
								Backgrnd	Color				
1	ICV	0.5	1	100	100		1.69	0.000	0.382	0.382	0.479	11/06/2019	12:28
2	ICB		1	100	100		1.98	0.000	0.000	0.000	0.000	11/06/2019	12:29
3	CCV1	0.5	1	100	100		1.69	0.000	0.375	0.375	0.470	11/06/2019	12:30
4	CCB1		1	100	100		2.10	0.000	0.000	0.000	0.000	11/06/2019	12:32
5	PB124567BL		1	2.51	100	7.22	1.98	0.000	0.000	0.000	0.000	11/06/2019	12:33
6	PB124567BS	20	1	2.52	100	7.26	1.89	0.000	0.376	0.376	0.471	11/06/2019	12:34
7	K5671-01		1	2.53	100	7.23	1.75	0.012	0.012	0.000	0.000	11/06/2019	12:35
8	K5671-01DUP		1	2.52	100	7.15	1.85	0.011	0.011	0.000	0.000	11/06/2019	12:36
9	K5671-01MSPre	40	2	2.51	100	7.42	1.69	0.015	0.380	0.365	0.458	11/06/2019	12:37
10	K5671-01MS2Ins	1284	40	2.52	100	7.16	1.98	0.011	0.654	0.643	0.806	11/06/2019	12:38
11	K5671-01MS3Post	40	2	2.54	100	7.26	1.75	0.012	0.390	0.378	0.474	11/06/2019	12:39
12	K5675-01		1	2.50	100	7.23	1.69	0.018	0.020	0.002	0.003	11/06/2019	12:40
13	CCV2	0.5	1	100	100		1.98	0.000	0.375	0.375	0.470	11/06/2019	12:41
14	CCB2		1	100	100		1.75	0.000	0.000	0.000	0.000	11/06/2019	12:42