

LB106136


7470A

INSTRUMENT ID:- CV1

Linear

 μ Abs.:

44353



A= 0.0000e+000

B= 2.2552e-004

C= -9.4839e-002

Rho= 0.9998174

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.064	-0.064	137	0.000	137				
0.2	0.200	0.178	-0.022	1211	0.0 %	1211				
2.5	2.500	2.526	0.026	11620	0.0 %	11620				
5.0	5.000	5.123	0.123	23139	0.0 %	23139				
7.5	7.500	7.529	0.029	33805	0.0 %	33805				
10.0	10.000	9.908	-0.092	44353	0.0 %	44353				

LB106136

Method: 7470A

Operator: Admin

Date of Analysis: 07 Nov 2019 08:28:20

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	30.0	137	-	0.0000	7470A	PPB	07 Nov 2019 08:37:04	S	Std
0.2 - 1	50.2	1211	-	0.2000	7470A	PPB	07 Nov 2019 08:39:12	S	Std
2.5 - 1	52.5	11620	-	2.5000	7470A	PPB	07 Nov 2019 08:41:21	S	Std
5.0 - 1	55.0	23139	-	5.0000	7470A	PPB	07 Nov 2019 08:43:33	S	Std
7.5 - 1	57.5	33805	-	7.5000	7470A	PPB	07 Nov 2019 08:45:41	S	Std
10.0 - 1	510.0	44353	-	10.0000	7470A	PPB	07 Nov 2019 08:47:48	S	Std
ICV87 - 1	ICV87	17095	3.7604	-	7470A	PPB	07 Nov 2019 08:51:36	U	SMPL
ICB87 - 1	ICB87	101	-0.0721	-	7470A	PPB	07 Nov 2019 08:53:42	U	SMPL
CCV85 - 1	CCV85	22019	4.8709	-	7470A	PPB	07 Nov 2019 08:55:48	U	SMPL
CCB85 - 1	CCB85	103	-0.0716	-	7470A	PPB	07 Nov 2019 08:57:54	U	SMPL
CRA - 1	CRA	1211	0.1783	-	7470A	PPB	07 Nov 2019 09:00:01	U	SMPL
HighStd - 1	HighStd	43688	9.7577	-	7470A	PPB	07 Nov 2019 09:02:06	U	SMPL
ChkStd - 1	ChkStd	30511	6.7860	-	7470A	PPB	07 Nov 2019 09:04:13	U	SMPL
PB124580BL - 1	PBW	100	-0.0723	-	7470A	PPB	07 Nov 2019 09:06:20	U	SMPL
PB124580BS - 1	LCSW	16076	3.5306	-	7470A	PPB	07 Nov 2019 09:08:28	U	SMPL
K5671-05 - 1	OR-02-110519	500	0.0179	-	7470A	PPB	07 Nov 2019 09:10:35	U	SMPL
K5671-05DUP - 1	OR-02-110519DUP	509	0.0200	-	7470A	PPB	07 Nov 2019 09:12:43	U	SMPL
K5671-05MS - 1	OR-02-110519MS	15527	3.4068	-	7470A	PPB	07 Nov 2019 09:14:51	U	SMPL
K5671-05MSD - 1	OR-02-110519MSD	16192	3.5568	-	7470A	PPB	07 Nov 2019 09:17:00	U	SMPL
K5675-04 - 1	TR-05-110519	454	0.0075	-	7470A	PPB	07 Nov 2019 09:19:06	U	SMPL
CCV86 - 1	CCV86	22039	4.8754	-	7470A	PPB	07 Nov 2019 09:21:12	U	SMPL
CCB86 - 1	CCB86	98	-0.0727	-	7470A	PPB	07 Nov 2019 09:23:19	U	SMPL
PB124580TB - 1	PB124559TB	457	0.0082	-	7470A	PPB	07 Nov 2019 09:25:25	U	SMPL
K5671-05LX5 - 1		118	-0.0682	-	7470A	PPB	07 Nov 2019 09:27:31	U	SMPL
K5671-05A - 1		16397	3.6030	-	7470A	PPB	07 Nov 2019 09:29:38	U	SMPL
CCV87 - 1	CCV87	22108	4.8910	-	7470A	PPB	07 Nov 2019 09:31:45	U	SMPL
CCB87 - 1	CCB87	104	-0.0714	-	7470A	PPB	07 Nov 2019 09:33:51	U	SMPL