

LB 10 6144


7471B

INSTRUMENT ID: CVI

Linear

μ Abs.:

58276



A= 0.0000e+000

B= 1.7175e-004

C= -3.0399e-002

Rho= 0.9999182

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.007	-0.007	139	0.000	139				
0.2	0.200	0.220	0.020	1458	0.0 %	1458				
2.5	2.500	2.513	0.013	14807	0.0 %	14807				
5.0	5.000	4.919	-0.081	28821	0.0 %	28821				
7.5	7.500	7.576	0.076	44290	0.0 %	44290				
10.0	10.000	9.978	-0.022	58276	0.0 %	58276				

LB106144 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 07 Nov 2019 10:50:51				
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0.0 - 1	50.0	139	-	0.0000	7471B	PPB	07 Nov 2019 10:51:37	S Std
0.2 - 1	50.2	1458	-	0.2000	7471B	PPB	07 Nov 2019 10:53:55	S Std
2.5 - 1	52.5	14807	-	2.5000	7471B	PPB	07 Nov 2019 10:56:13	S Std
5.0 - 1	55.0	28821	-	5.0000	7471B	PPB	07 Nov 2019 10:58:31	S Std
7.5 - 1	57.5	44290	-	7.5000	7471B	PPB	07 Nov 2019 11:00:48	S Std
10.0 - 1	510.0	58276	-	10.0000	7471B	PPB	07 Nov 2019 11:03:04	S Std
ICV88 - 1	ICV88	23324	3.9754	-	7471B	PPB	07 Nov 2019 11:08:28	U SMPL
ICB88 - 1	ICB88	88	-0.0153	-	7471B	PPB	07 Nov 2019 11:10:43	U SMPL
CCV88 - 1	CCV88	28645	4.8892	-	7471B	PPB	07 Nov 2019 11:13:00	U SMPL
CCB88 - 1	CCB88	55	-0.0210	-	7471B	PPB	07 Nov 2019 11:15:15	U SMPL
CRA - 1	CRA	1444	0.2176	-	7471B	PPB	07 Nov 2019 11:17:31	U SMPL
HighStd - 1	HighStd	58462	10.0102	-	7471B	PPB	07 Nov 2019 11:19:46	U SMPL
ChkStd - 1	ChkStd	40524	6.9294	-	7471B	PPB	07 Nov 2019 11:22:02	U SMPL
PB124585BL - 1	PBS	74	-0.0177	-	7471B	PPB	07 Nov 2019 11:24:19	U SMPL
PB124585BS - 1	LCSS	22981	3.9165	-	7471B	PPB	07 Nov 2019 11:26:36	U SMPL
K5671-01 - 1	OR-02-110519	5267	0.8742	-	7471B	PPB	07 Nov 2019 11:37:53	U SMPL
K5671-01DUP - 1	OR-02-110519DUP	4949	0.8196	-	7471B	PPB	07 Nov 2019 11:40:10	U SMPL
K5671-01MS - 1	OR-02-110519MS	24249	4.1343	-	7471B	PPB	07 Nov 2019 11:42:28	U SMPL
K5671-01MSD - 1	OR-02-110519MSD	24151	4.1174	-	7471B	PPB	07 Nov 2019 11:44:45	U SMPL
K5675-01 - 1	TR-05-110519	5072	0.8407	-	7471B	PPB	07 Nov 2019 11:47:01	U SMPL
CCV89 - 1	CCV89	29038	4.9567	-	7471B	PPB	07 Nov 2019 11:49:17	U SMPL
CCB89 - 1	CCB89	-14	-0.0328	-	7471B	PPB	07 Nov 2019 11:51:33	U SMPL
K5701-01 - 1	J-1	7683	1.2891	-	7471B	PPB	07 Nov 2019 11:53:49	U SMPL
K5702-01 - 1	H-1	2980	0.4814	-	7471B	PPB	07 Nov 2019 11:56:06	U SMPL
K5710-03 - 1	TP- 31-S	1318	0.1960	-	7471B	PPB	07 Nov 2019 11:58:30	U SMPL
K5710-07 - 1	TP- 32-S	1149	0.1669	-	7471B	PPB	07 Nov 2019 12:00:49	U SMPL
K5710-11 - 1	TP- 32-D	681	0.0866	-	7471B	PPB	07 Nov 2019 12:03:05	U SMPL
K5712-01 - 1	SB-2A	27847	4.7522	-	7471B	PPB	07 Nov 2019 12:05:22	U SMPL
K5712-02 - 1	SB-3A	89442	15.3308	-	7471B	PPB	07 Nov 2019 12:07:39	U SMPL
K5671-01LX5 - 1		942	0.1314	-	7471B	PPB	07 Nov 2019 12:09:56	U SMPL
K5671-01A - 1		24313	4.1452	-	7471B	PPB	07 Nov 2019 12:12:15	U SMPL
CCV90 - 1	CCV90	29366	5.0131	-	7471B	PPB	07 Nov 2019 12:14:31	U SMPL
CCB90 - 1	CCB90	-8	-0.0318	-	7471B	PPB	07 Nov 2019 12:16:47	U SMPL
K5712-02DLX5 - 1		17910	3.0456	-	7471B	PPB	07 Nov 2019 12:19:04	U SMPL
CCV91 - 1	CCV91	29005	4.9511	-	7471B	PPB	07 Nov 2019 12:21:21	U SMPL
CCB91 - 1	CCB91	0	-0.0304	-	7471B	PPB	07 Nov 2019 12:23:37	U SMPL