


ISM02.4

INSTRUMENT ID :- CVI
LB106150

Linear

 μ Abs.:

44623



A= 0.0000e+000

B= 2.2360e-004 slope

C= -2.0194e-002 y-intercept

Rho= 0.9999477

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.009	-0.009	51	0.000	51					
0.05	0.050										-4.5
0.20	0.200	0.191	-0.009	946	0.0 %	946					-1.2
2.50	2.500	2.470	-0.030	11135	0.0 %	11135					1.48
5.00	5.000	5.074	0.074	22780	0.0 %	22780					0
7.50	7.500	7.517	0.017	33706	0.0 %	33706					0
10.0	10.000	9.958	-0.042	44623	0.0 %	44623					

LB106150
INSTRUMENT ID : CV1

Sample ID	Extended II μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0 S0.0	51	-		0 ISM02.4	PPB	11/7/2019 13:31	Std
0.2 S0.2	946	-		0.2 ISM02.4	PPB	11/7/2019 13:33	Std
2.5 S2.5	11135	-		2.5 ISM02.4	PPB	11/7/2019 13:36	Std
5 S5.0	22780	-		5 ISM02.4	PPB	11/7/2019 13:38	Std
7.5 S7.5	33706	-		7.5 ISM02.4	PPB	11/7/2019 13:40	Std
10 S10.0	44623	-		10 ISM02.4	PPB	11/7/2019 13:43	Std
ICV	ICV	18454	4.1062 -	ISM02.4	PPB	11/7/2019 13:49	SMPL
ICB	ICB	1	-0.02 -	ISM02.4	PPB	11/7/2019 13:51	SMPL
CCV092	CCV092	23162	5.1589 -	ISM02.4	PPB	11/7/2019 13:59	SMPL
CCB092	CCB092	151	0.0136 -	ISM02.4	PPB	11/7/2019 14:01	SMPL
PB124577BL	PBS577	524	0.097 -	ISM02.4	PPB	11/7/2019 14:03	SMPL
K5677-01	MJLM53	2011	0.4295 -	ISM02.4	PPB	11/7/2019 14:06	SMPL
K5677-02	MJLM57	2499	0.5386 -	ISM02.4	PPB	11/7/2019 14:08	SMPL
K5677-03	MJLM61	3056	0.6631 -	ISM02.4	PPB	11/7/2019 14:10	SMPL
K5677-04	MJLM65	2598	0.5607 -	ISM02.4	PPB	11/7/2019 14:12	SMPL
K5677-05	MJLM69	15666	3.4828 -	ISM02.4	PPB	11/7/2019 14:15	SMPL
K5677-06	MJLM73	2734	0.5911 -	ISM02.4	PPB	11/7/2019 14:17	SMPL
K5677-07	MJLM77	1931	0.4116 -	ISM02.4	PPB	11/7/2019 14:19	SMPL
K5677-08	MJLM81	5398	1.1868 -	ISM02.4	PPB	11/7/2019 14:21	SMPL
K5677-09	MJLM85	4979	1.0931 -	ISM02.4	PPB	11/7/2019 14:24	SMPL
K5677-10	MJLM89	2011	0.4295 -	ISM02.4	PPB	11/7/2019 14:26	SMPL
K5677-11	MJLM93	4877	1.0703 -	ISM02.4	PPB	11/7/2019 14:28	SMPL
K5677-12	MJLM97	3441	0.7492 -	ISM02.4	PPB	11/7/2019 14:31	SMPL
K5677-13	MJLM98	12144	2.6953 -	ISM02.4	PPB	11/7/2019 14:33	SMPL
K5677-14	MJLM99	5055	1.1101 -	ISM02.4	PPB	11/7/2019 14:35	SMPL
K5677-15	MJLMA0	4647	1.0189 -	ISM02.4	PPB	11/7/2019 14:37	SMPL
K5677-16	MJLMA1	3994	0.8729 -	ISM02.4	PPB	11/7/2019 14:50	SMPL
K5677-17	MJLMA1D	3916	0.8554 -	ISM02.4	PPB	11/7/2019 14:52	SMPL
K5677-18	MJLMA1S	15341	3.4101 -	ISM02.4	PPB	11/7/2019 14:54	SMPL
K5677-19	MJLMK4	3827	0.8355 -	ISM02.4	PPB	11/7/2019 14:57	SMPL
CCV093	CCV093	23190	5.1652 -	ISM02.4	PPB	11/7/2019 14:59	SMPL
CCB093	CCB093	-12	-0.0229 -	ISM02.4	PPB	11/7/2019 15:01	SMPL
K5677-20	MJLMK6	3731	0.8141 -	ISM02.4	PPB	11/7/2019 15:04	SMPL
K5677-21	MJLMM1	37	-0.0119 -	ISM02.4	PPB	11/7/2019 15:06	SMPL
PB124607BL	PBS607	-18	-0.0242 -	ISM02.4	PPB	11/7/2019 15:08	SMPL
K5638-19	MX2J61	10703	2.373 -	ISM02.4	PPB	11/7/2019 15:10	SMPL
K5638-20	MX2J61D	12036	2.6711 -	ISM02.4	PPB	11/7/2019 15:13	SMPL
K5638-21	MX2J61S	24682	5.4988 -	ISM02.4	PPB	11/7/2019 15:15	SMPL
CCV094	CCV094	23302	5.1902 -	ISM02.4	PPB	11/7/2019 15:17	SMPL
CCB094	CCB094	-7	-0.0218 -	ISM02.4	PPB	11/7/2019 15:19	SMPL