



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB106153

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	11/07/2019	12:25
2	CAL2	1	Water	NA	NA	20.1	7.01	11/07/2019	12:26
3	CAL3	1	Water	NA	NA	20.1	10.03	11/07/2019	12:27
4	ICV	1	Water	NA	NA	20.2	6.99	11/07/2019	12:30
5	CCV1	1	Water	NA	NA	20.1	2.02	11/07/2019	12:32
6	K5730-01	1	Solid	20.03	20	20.3	7.00	11/07/2019	12:34
7	K5730-01DUP	1	Solid	20.04	20	20.4	7.01	11/07/2019	12:35
8	CCV2	1	Water	NA	NA	20.1	12.01	11/07/2019	12:36

LB 106153

WORKLIST(Hardcopy Internal Chain)

Date : 11-07-2019 12:16:43

Department : Wet-Chemistry

WorkList ID : 133009

WorkList Name : PH K5730

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/06/2019	Solid	K5730-01-17	pH	Cool 4 deg C	H0ME01	H61	HILTON-STOCKPILE	11/06/2019	9045D

Date/Time 11-07-19 12:20
Received by: Jo
Relinquished by: cl

Date/Time 11-07-19 13:15
Received by: cl
Relinquished by: Jo