


7471B

LB106168
INSTRUMENT ID:- W2 μ Abs.:

59712



Linear

A= 0.0000e+000

B= 1.6698e-004

C= 1.2706e-002

Rho= 0.9999457

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.037	0.037	148	0.000	148				
0.2	0.200	0.207	0.007	1163	0.0 %	1163				
2.5	2.500	2.430	-0.070	14477	0.0 %	14477				
5.0	5.000	4.995	-0.005	29834	0.0 %	29834				
7.5	7.500	7.547	0.047	45122	0.0 %	45122				
10.0	10.000	9.984	-0.016	59712	0.0 %	59712				

LB106168 INSTRUMENT ID CV1

Method: 7471B		Operator: Admin		Date of Analysis: 08 Nov 2019 09:48:36					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	500	148	-	0.0000	7471B	PPB	08 Nov 2019 09:58:51	S	Std
0.2 - 1	602	1163	-	0.2000	7471B	PPB	08 Nov 2019 10:01:08	S	Std
2.5 - 1	505	14477	-	2.5000	7471B	PPB	08 Nov 2019 10:03:26	S	Std
5.0 - 1	550	29834	-	5.0000	7471B	PPB	08 Nov 2019 10:05:43	S	Std
7.5 - 1	575	45122	-	7.5000	7471B	PPB	08 Nov 2019 10:08:00	S	Std
10.0 - 1	5100	59712	-	10.0000	7471B	PPB	08 Nov 2019 10:10:16	S	Std
ICV90 - 1	ICV90	23234	3.8924	-	7471B	PPB	08 Nov 2019 10:15:38	U	SMPL
ICB90 - 1	ICB90	119	0.0326	-	7471B	PPB	08 Nov 2019 10:17:53	U	SMPL
CCV01 - 1	CCV01	30054	5.0312	-	7471B	PPB	08 Nov 2019 10:20:09	U	SMPL
CCB01 - 1	CCB01	108	0.0307	-	7471B	PPB	08 Nov 2019 10:22:24	U	SMPL
CRA - 1	CRA	1183	0.2102	-	7471B	PPB	08 Nov 2019 10:24:39	U	SMPL
HighStd - 1	HighStd	60192	10.0638	-	7471B	PPB	08 Nov 2019 10:26:54	U	SMPL
ChkStd - 1	ChkStd	41732	6.9813	-	7471B	PPB	08 Nov 2019 10:29:10	U	SMPL
PB124629BL - 1	PBS	85	0.0269	-	7471B	PPB	08 Nov 2019 10:31:27	U	SMPL
PB124629BS - 1	LCSS	22046	3.6940	-	7471B	PPB	08 Nov 2019 10:33:43	U	SMPL
K5730-01 - 1	HILTON-STOCKPILE	4997	0.8471	-	7471B	PPB	08 Nov 2019 10:36:00	U	SMPL
K5742-01 - 1	SB-10-(4-6)	2273	0.3923	-	7471B	PPB	08 Nov 2019 10:38:17	U	SMPL
K5742-02 - 1	SB-11-(12-14)	254	0.0551	-	7471B	PPB	08 Nov 2019 10:40:35	U	SMPL
K5742-03 - 1	SB-17-(6-8)	185	0.0436	-	7471B	PPB	08 Nov 2019 10:42:52	U	SMPL
K5742-04 - 1	SB-18-(12-14)	664	0.1236	-	7471B	PPB	08 Nov 2019 10:45:08	U	SMPL
CCV02 - 1	CCV02	29945	5.0130	-	7471B	PPB	08 Nov 2019 10:47:23	U	SMPL
CCB02 - 1	CCB02	-32	0.0074	-	7471B	PPB	08 Nov 2019 10:49:39	U	SMPL
K5742-05 - 1	SB-22-(14-15)	482	0.0932	-	7471B	PPB	08 Nov 2019 10:51:55	U	SMPL
K5742-06 - 1	SB-30-(6-8)	199	0.0459	-	7471B	PPB	08 Nov 2019 10:54:11	U	SMPL
K5742-07 - 1	SB-35-(6-8)	1142	0.2034	-	7471B	PPB	08 Nov 2019 10:56:26	U	SMPL
K5742-08 - 1	SB-26-(6-8)	596	0.1122	-	7471B	PPB	08 Nov 2019 10:58:42	U	SMPL
K5742-09 - 1	SB-37-(8-10)	2693	0.4624	-	7471B	PPB	08 Nov 2019 11:00:58	U	SMPL
K5742-10 - 1	SB-38-(2-4)	699	0.1294	-	7471B	PPB	08 Nov 2019 11:03:15	U	SMPL
K5742-11 - 1	K5742-10MS	23204	3.8874	-	7471B	PPB	08 Nov 2019 11:05:31	U	SMPL
K5742-12 - 1	K5742-10MSD	22943	3.8438	-	7471B	PPB	08 Nov 2019 11:07:50	U	SMPL
K5746-05 - 1	TP-1A-COMP	3553	0.6060	-	7471B	PPB	08 Nov 2019 11:10:08	U	SMPL
PB124627BL - 1	PBS	13	0.0149	-	7471B	PPB	08 Nov 2019 11:12:24	U	SMPL
CCV03 - 1	CCV03	29763	4.9827	-	7471B	PPB	08 Nov 2019 11:14:39	U	SMPL
CCB03 - 1	CCB03	-15	0.0102	-	7471B	PPB	08 Nov 2019 11:16:55	U	SMPL
PB124627BS - 1	LCSS	21825	3.6571	-	7471B	PPB	08 Nov 2019 11:19:11	U	SMPL
K5722-01 - 1	4-A	4109	0.6988	-	7471B	PPB	08 Nov 2019 11:21:27	U	SMPL
K5722-01MS - 1	4-AMS	24140	4.0437	-	7471B	PPB	08 Nov 2019 11:26:00	U	SMPL
K5722-01MSD - 1	4-AMSD	27581	4.6183	-	7471B	PPB	08 Nov 2019 11:28:16	U	SMPL
K5722-02 - 1	4-B	6021	1.0181	-	7471B	PPB	08 Nov 2019 11:30:33	U	SMPL
K5722-03 - 1	4-C	3957	0.6735	-	7471B	PPB	08 Nov 2019 11:32:51	U	SMPL
K5722-04 - 1	4-D	7014	1.1839	-	7471B	PPB	08 Nov 2019 11:35:08	U	SMPL
K5722-05 - 1	5-A	6239	1.0545	-	7471B	PPB	08 Nov 2019 11:37:26	U	SMPL
K5722-06 - 1	5-B	5797	0.9807	-	7471B	PPB	08 Nov 2019 11:39:42	U	SMPL
CCV04 - 1	CCV04	29950	5.0139	-	7471B	PPB	08 Nov 2019 11:42:00	U	SMPL
CCB04 - 1	CCB04	-14	0.0104	-	7471B	PPB	08 Nov 2019 11:44:16	U	SMPL
K5722-07 - 1	5-C	9710	1.6341	-	7471B	PPB	08 Nov 2019 11:46:32	U	SMPL
K5722-08 - 1	6-A	8802	1.4825	-	7471B	PPB	08 Nov 2019 11:48:47	U	SMPL
K5722-09 - 1	6-B	6048	1.0226	-	7471B	PPB	08 Nov 2019 11:51:04	U	SMPL
K5722-10 - 1	6-C	5687	0.9623	-	7471B	PPB	08 Nov 2019 11:53:19	U	SMPL
K5722-11 - 1	1-A	7743	1.3057	-	7471B	PPB	08 Nov 2019 11:55:36	U	SMPL
K5722-12 - 1	1-B	6585	1.1123	-	7471B	PPB	08 Nov 2019 11:57:53	U	SMPL
K5722-13 - 1	1-C	3003	0.5142	-	7471B	PPB	08 Nov 2019 12:00:10	U	SMPL
K5722-14 - 1	2-A	62512	10.4512	-	7471B	PPB	08 Nov 2019 12:02:28	U	SMPL
K5722-15 - 1	2-B	52420	8.7660	-	7471B	PPB	08 Nov 2019 12:04:45	U	SMPL
K5722-16 - 1	2-C	56925	9.5183	-	7471B	PPB	08 Nov 2019 12:07:02	U	SMPL
CCV05 - 1	CCV05	29984	5.0196	-	7471B	PPB	08 Nov 2019 12:09:20	U	SMPL
CCB05 - 1	CCB05	-32	0.0074	-	7471B	PPB	08 Nov 2019 12:11:36	U	SMPL
K5722-17 - 1	3-A	8111	1.3671	-	7471B	PPB	08 Nov 2019 12:13:53	U	SMPL
K5722-18 - 1	3-B	7848	1.3232	-	7471B	PPB	08 Nov 2019 12:16:08	U	SMPL
K5722-19 - 1	3-C	9106	1.5333	-	7471B	PPB	08 Nov 2019 12:18:24	U	SMPL
K5722-20 - 1	3-D	6273	1.0602	-	7471B	PPB	08 Nov 2019 12:20:40	U	SMPL
PB124625BL - 1	PBS	13	0.0149	-	7471B	PPB	08 Nov 2019 12:22:57	U	SMPL
PB124625BS - 1	LCSS	21981	3.6832	-	7471B	PPB	08 Nov 2019 12:25:13	U	SMPL
K5668-01 - 1	HD-01-110619	2916	0.4996	-	7471B	PPB	08 Nov 2019 12:27:30	U	SMPL
K5668-01DUP - 1	HD-01-110619DUP	2051	0.3552	-	7471B	PPB	08 Nov 2019 12:29:47	U	SMPL
K5668-01MS - 1	HD-01-110619MS	24576	4.1165	-	7471B	PPB	08 Nov 2019 12:32:04	U	SMPL
K5668-01MSD - 1	HD-01-110619MSD	25085	4.2015	-	7471B	PPB	08 Nov 2019 12:34:20	U	SMPL
CCV06 - 1	CCV06	30134	5.0446	-	7471B	PPB	08 Nov 2019 12:36:37	U	SMPL
CCB06 - 1	CCB06	-7	0.0115	-	7471B	PPB	08 Nov 2019 12:38:54	U	SMPL
K5675-06 - 1	TR-04-110619	2497	0.4297	-	7471B	PPB	08 Nov 2019 12:41:11	U	SMPL
K5676-06 - 1	SU-04-110619	6829	1.1530	-	7471B	PPB	08 Nov 2019 12:43:28	U	SMPL
K5717-01 - 1	VC-TP	319	0.0660	-	7471B	PPB	08 Nov 2019 12:45:45	U	SMPL
K5723-01 - 1	7-A	15629	2.6225	-	7471B	PPB	08 Nov 2019 12:48:01	U	SMPL
K5723-02 - 1	7-B	1801	0.3134	-	7471B	PPB	08 Nov 2019 12:50:18	U	SMPL
K5723-03 - 1	7-C	5990	1.0129	-	7471B	PPB	08 Nov 2019 12:52:35	U	SMPL
K5723-04 - 1	8-A	10353	1.7415	-	7471B	PPB	08 Nov 2019 12:54:51	U	SMPL
K5723-05 - 1	8-B	8203	1.3825	-	7471B	PPB	08 Nov 2019 12:57:09	U	SMPL
K5723-06 - 1	8-C	6476	1.0941	-	7471B	PPB	08 Nov 2019 12:59:26	U	SMPL
K5723-07 - 1	8-D	7192	1.2137	-	7471B	PPB	08 Nov 2019 13:01:43	U	SMPL
CCV07 - 1	CCV07	29527	4.9432	-	7471B	PPB	08 Nov 2019 13:03:59	U	SMPL
CCB07 - 1	CCB07	6	0.0137	-	7471B	PPB	08 Nov 2019 13:06:16	U	SMPL

LB106168 INSTRUMENT ID CV1

Method: 7471B

Operator: Admin

Date of Analysis: 08 Nov 2019 09:48:36

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
K5723-08 - 1	9-A	273	0.0583		-7471B	PPB	08 Nov 2019 13:08:33	U	SMPL
K5723-09 - 1	9-B	246	0.0538		-7471B	PPB	08 Nov 2019 13:10:50	U	SMPL
K5723-10 - 1	9-C	283	0.0600		-7471B	PPB	08 Nov 2019 13:13:07	U	SMPL
K5723-11 - 1	9-D	286	0.0605		-7471B	PPB	08 Nov 2019 13:15:23	U	SMPL
K5723-12 - 1	10-A	623	0.1167		-7471B	PPB	08 Nov 2019 13:17:40	U	SMPL
K5723-13 - 1	10-B	167	0.0406		-7471B	PPB	08 Nov 2019 13:19:56	U	SMPL
K5723-14 - 1	10-C	182	0.0431		-7471B	PPB	08 Nov 2019 13:22:13	U	SMPL
K5723-15 - 1	10-D	212	0.0481		-7471B	PPB	08 Nov 2019 13:24:30	U	SMPL
K5723-16 - 1	4-E	5362	0.9081		-7471B	PPB	08 Nov 2019 13:26:47	U	SMPL
K5742-10DUP - 1	SB-38-(2-4)DUP	706	0.1306		-7471B	PPB	08 Nov 2019 13:29:04	U	SMPL
CCV08 - 1	CCV08	29678	4.9685		-7471B	PPB	08 Nov 2019 13:31:20	U	SMPL
CCB08 - 1	CCB08	7	0.0139		-7471B	PPB	08 Nov 2019 13:33:37	U	SMPL
K5742-10LX5 - 1		138	0.0357		-7471B	PPB	08 Nov 2019 13:35:54	U	SMPL
K5742-10A - 1		23047	3.8612		-7471B	PPB	08 Nov 2019 13:38:11	U	SMPL
K5722-01LX5 - 1		956	0.1723		-7471B	PPB	08 Nov 2019 13:40:28	U	SMPL
K5722-01A - 1		24967	4.1818		-7471B	PPB	08 Nov 2019 13:42:45	U	SMPL
K5668-01LX5 - 1		429	0.0843		-7471B	PPB	08 Nov 2019 13:45:01	U	SMPL
K5668-01A - 1		24022	4.0240		-7471B	PPB	08 Nov 2019 13:47:18	U	SMPL
K5722-01DUP - 1	4-ADUP	4031	0.6858		-7471B	PPB	08 Nov 2019 13:49:35	U	SMPL
K5722-14DLX2 - 1		30543	5.1129		-7471B	PPB	08 Nov 2019 13:51:52	U	SMPL
CCV09 - 1	CCV09	30194	5.0546		-7471B	PPB	08 Nov 2019 13:54:09	U	SMPL
CCB09 - 1	CCB09	15	0.0152		-7471B	PPB	08 Nov 2019 13:56:27	U	SMPL