

LB106174


7470A

INSTRUMENT ID : CU1

Linear

 μ Abs.:

4484E



A= 0.0000e+000

B= 2.2248e-004

C= -4.9168e-002

Rho= 0.9998896

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.040	-0.040	43	0.000	43				
0.2	0.200	0.224	0.024	1229	0.0 %	1229				
2.5	2.500	2.456	-0.044	11258	0.0 %	11258				
5.0	5.000	5.072	0.072	23020	0.0 %	23020				
7.5	7.500	7.559	0.059	34198	0.0 %	34198				
10.0	10.000	9.928	-0.072	44846	0.0 %	44846				

LB106174 INSTRUMENT ID CV1

Method: 7470A		Operator: Admin		Date of Analysis: 08 Nov 2019 14:32:00					
Sample ID	Extended ID	µ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1		43	-	0.0000	7470A	PPB	08 Nov 2019 14:38:02	S	Std
0.2 - 1		1229	-	0.2000	7470A	PPB	08 Nov 2019 14:40:11	S	Std
2.5 - 1		11258	-	2.5000	7470A	PPB	08 Nov 2019 14:42:20	S	Std
5.0 - 1		23020	-	5.0000	7470A	PPB	08 Nov 2019 14:44:28	S	Std
7.5 - 1		34198	-	7.5000	7470A	PPB	08 Nov 2019 14:46:36	S	Std
10.0 - 1		44846	-	10.0000	7470A	PPB	08 Nov 2019 14:48:43	S	Std
ICV91 - 1	ICV91	17893	3.9317	-	7470A	PPB	08 Nov 2019 15:00:38	U	SMPL
ICB91 - 1	ICB91	0	-0.0492	-	7470A	PPB	08 Nov 2019 15:02:44	U	SMPL
CCV10 - 1	CCV10	23253	5.1242	-	7470A	PPB	08 Nov 2019 15:04:51	U	SMPL
CCB10 - 1	CCB10	-1	-0.0494	-	7470A	PPB	08 Nov 2019 15:06:57	U	SMPL
CRA - 1	CRA	1228	0.2240	-	7470A	PPB	08 Nov 2019 15:09:03	U	SMPL
HighStd - 1	HighStd	45146	9.9950	-	7470A	PPB	08 Nov 2019 15:11:10	U	SMPL
ChkStd - 1	ChkStd	31097	6.8693	-	7470A	PPB	08 Nov 2019 15:13:16	U	SMPL
PB124635BL - 1	PBW	-19	-0.0534	-	7470A	PPB	08 Nov 2019 15:15:23	U	SMPL
PB124635BS - 1	LCSW	16637	3.6523	-	7470A	PPB	08 Nov 2019 15:17:30	U	SMPL
K5668-04 - 1	HD-01-110619	13	-0.0463	-	7470A	PPB	08 Nov 2019 15:19:38	U	SMPL
K5668-04DUP - 1	HD-01-110619DUP	34	-0.0416	-	7470A	PPB	08 Nov 2019 15:21:45	U	SMPL
K5668-04MS - 1	HD-01-110619MS	15401	3.3773	-	7470A	PPB	08 Nov 2019 15:23:53	U	SMPL
K5668-04MSD - 1	HD-01-110619MSD	15709	3.4458	-	7470A	PPB	08 Nov 2019 15:26:02	U	SMPL
K5675-08 - 1	TR-04-110619	1	-0.0489	-	7470A	PPB	08 Nov 2019 15:28:08	U	SMPL
CCV11 - 1	CCV11	23233	5.1197	-	7470A	PPB	08 Nov 2019 15:30:14	U	SMPL
CCB11 - 1	CCB11	3	-0.0485	-	7470A	PPB	08 Nov 2019 15:32:20	U	SMPL
K5676-10 - 1	SU-04-110619	34	-0.0416	-	7470A	PPB	08 Nov 2019 15:34:27	U	SMPL
K5710-04 - 1	TP- 31-S	128	-0.0207	-	7470A	PPB	08 Nov 2019 15:36:34	U	SMPL
K5710-08 - 1	TP- 32-S	105	-0.0258	-	7470A	PPB	08 Nov 2019 15:38:41	U	SMPL
K5710-12 - 1	TP- 32-D	47	-0.0387	-	7470A	PPB	08 Nov 2019 15:40:48	U	SMPL
K5717-02 - 1	VC-TP	33	-0.0418	-	7470A	PPB	08 Nov 2019 15:42:55	U	SMPL
PB124636BL - 1	PBW	638	0.0928	-	7470A	PPB	08 Nov 2019 15:45:04	U	SMPL
PB124636BS - 1	LCSW	16797	3.6878	-	7470A	PPB	08 Nov 2019 15:47:13	U	SMPL
K5683-02 - 1	OILY-WATER	828	0.1350	-	7470A	PPB	08 Nov 2019 15:49:21	U	SMPL
K5683-02DUP - 1	OILY-WATERDUP	809	0.1308	-	7470A	PPB	08 Nov 2019 15:51:30	U	SMPL
K5683-02MS - 1	OILY-WATERMS	13390	2.9299	-	7470A	PPB	08 Nov 2019 15:53:37	U	SMPL
CCV12 - 1	CCV12	22796	5.0225	-	7470A	PPB	08 Nov 2019 15:55:44	U	SMPL
CCB12 - 1	CCB12	-35	-0.0570	-	7470A	PPB	08 Nov 2019 15:57:52	U	SMPL
K5683-02MSD - 1	OILY-WATERMSD	13473	2.9483	-	7470A	PPB	08 Nov 2019 16:00:00	U	SMPL
K5703-01 - 1	OILY-WATER	69277	15.3637	-	7470A	PPB	08 Nov 2019 16:02:08	U	SMPL
K5742-13 - 1	(ERB)-EQUIPMENT-RINSATE-BLANK	-66	-0.0639	-	7470A	PPB	08 Nov 2019 16:04:17	U	SMPL
PB124661BL - 1	PBW	-6	-0.0505	-	7470A	PPB	08 Nov 2019 16:06:26	U	SMPL
PB124661BS - 1	LCSW	16326	3.5831	-	7470A	PPB	08 Nov 2019 16:08:35	U	SMPL
K5683-03 - 1	OILY-WATER	4	-0.0483	-	7470A	PPB	08 Nov 2019 16:10:41	U	SMPL
K5683-03DUP - 1	OILY-WATERDUP	35	-0.0414	-	7470A	PPB	08 Nov 2019 16:12:50	U	SMPL
K5683-03MS - 1	OILY-WATERMS	13909	3.0453	-	7470A	PPB	08 Nov 2019 16:14:58	U	SMPL
K5683-03MSD - 1	OILY-WATERMSD	13952	3.0549	-	7470A	PPB	08 Nov 2019 16:17:07	U	SMPL
PB124635TB - 1	PB124578TB	-18	-0.0532	-	7470A	PPB	08 Nov 2019 16:19:15	U	SMPL
CCV13 - 1	CCV13	22557	4.9693	-	7470A	PPB	08 Nov 2019 16:21:24	U	SMPL
CCB13 - 1	CCB13	459	0.0530	-	7470A	PPB	08 Nov 2019 16:23:31	U	SMPL
PB124661TB - 1	PB124605TB	4	-0.0483	-	7470A	PPB	08 Nov 2019 16:25:39	U	SMPL
K5668-04LX5 - 1		-1	-0.0494	-	7470A	PPB	08 Nov 2019 16:27:46	U	SMPL
K5668-04A - 1		15882	3.4843	-	7470A	PPB	08 Nov 2019 16:29:54	U	SMPL
K5683-02LX5 - 1		162	-0.0131	-	7470A	PPB	08 Nov 2019 16:32:03	U	SMPL
K5683-03LX5 - 1		-51	-0.0605	-	7470A	PPB	08 Nov 2019 16:36:19	U	SMPL
K5703-01DLX5 - 1		14138	3.0963	-	7470A	PPB	08 Nov 2019 16:40:39	U	SMPL
K5683-02A - 1		16796	3.6876	-	7470A	PPB	08 Nov 2019 16:57:59	U	SMPL
K5683-03A - 1		16691	3.6643	-	7470A	PPB	08 Nov 2019 17:00:06	U	SMPL
CCV14 - 1	CCV14	23565	5.1936	-	7470A	PPB	08 Nov 2019 17:02:14	U	SMPL
CCB14 - 1	CCB14	275	0.0120	-	7470A	PPB	08 Nov 2019 17:04:23	U	SMPL