



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB106175

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	11/08/2019	13:45
2	CAL2	1	Water	NA	NA	20.1	7.01	11/08/2019	13:46
3	CAL3	1	Water	NA	NA	20.1	10.03	11/08/2019	13:47
4	ICV	1	Water	NA	NA	20.2	7.00	11/08/2019	13:48
5	CCV1	1	Water	NA	NA	20.1	2.01	11/08/2019	13:50
6	K5765-02	1	Water	NA	NA	13.4	7.51	11/08/2019	13:51
7	K5765-02DUP	1	Water	NA	NA	13.5	7.52	11/08/2019	14:00
8	K5765-03	1	Water	NA	NA	12.8	6.95	11/08/2019	14:01
9	K5765-04	1	Water	NA	NA	14.2	7.16	11/08/2019	14:05
10	CCV2	1	Water	NA	NA	20.2	12.02	11/08/2019	14:10

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K5765

WorkList ID : 133055

Department : Wet-Chemistry

Date : 11-08-2019 13:36:16

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/07/2019	Water	K5765-02 ~ 5	pH	Cool 4 deg C	TETR06	H61	TT-009-IDWGW-20191107	11/07/2019	9040C
11/07/2019	Water	K5765-03 ~ 5	pH	Cool 4 deg C	TETR06	H61	TT-010-IDWGW-20191107	11/07/2019	9040C
11/07/2019	Water	K5765-04 ~ 5	pH	Cool 4 deg C	TETR06	H61	TT-011-IDWGW-20191107	11/07/2019	9040C

13:40 J0

Date/Time 11-08-19 14:00
 Received by: J0
 Relinquished by: es

Date/Time 11-08-19 14:35
 Received by: es
 Relinquished by: J0