



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB106176

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	11/08/2019	13:45
2	CAL2	1	Water	NA	NA	20.1	7.01	11/08/2019	13:46
3	CAL3	1	Water	NA	NA	20.1	10.03	11/08/2019	13:47
4	ICV	1	Water	NA	NA	20.2	7.00	11/08/2019	13:48
5	CCV1	1	Water	NA	NA	20.2	2.01	11/08/2019	14:15
6	K5773-01	1	Water	NA	NA	18.5	5.05	11/08/2019	14:18
7	K5773-01DUP	1	Water	NA	NA	18.6	5.06	11/08/2019	14:19
8	K5773-02	1	Water	NA	NA	18.1	4.96	11/08/2019	14:20
9	K5773-03	1	Water	NA	NA	18.7	5.00	11/08/2019	14:22
10	CCV2	1	Water	NA	NA	20.1	12.02	11/08/2019	14:25

13106176

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K5773 WorkList ID : 133056 Department : Wet-Chemistry Date : 11-08-2019 13:36:48

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/07/2019	Water	K5773-01	pH	Cool 4 deg C	STER13	H51	CELL-A-POOL-WATER	11/07/2019	SM 4500-H B
11/07/2019	Water	K5773-02	pH	Cool 4 deg C	STER13	H51	CELL-B-POOL-WATER	11/07/2019	SM 4500-H B
11/07/2019	Water	K5773-03	pH	Cool 4 deg C	STER13	H51	CELL-C-POOL-WATER	11/07/2019	SM 4500-H B

Date/Time 11-08-19 13:40
 Received by: Ja
 Relinquished by: Ja

Date/Time 11-08-19 15:00
 Received by: Ja
 Relinquished by: Ja