



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB106192

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	11/11/2019	09:35
2	K5777-07		1	9	100.00	>212.0	>212.0	11/11/2019	09:55
3	K5777-07DUP		1	9	100.00	>212.0	>212.0	11/11/2019	10:20
4	K5777-08		1	8	100.00	>212.0	>212.0	11/11/2019	10:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB106192

WorkList Name : FP111119 WorkList ID : 133110 Department : Wet-Chemistry Date : 11-11-2019 09:01:37

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/15/2019	Water	K5777-07	Flash Point	Cool 4 deg C	TULL02	H61	AOC-7-MIDDLE	11/08/2019	1010A
11/15/2019	Water	K5777-08	Flash Point	Cool 4 deg C	TULL02	H61	AOC-7-NORTH	11/08/2019	1010A

Date/Time 11/11/19 09:05
 Received by: KS
 Relinquished by: JP

Date/Time 11/11/19 10:50
 Received by: JP
 Relinquished by: KS