



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB106207

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	3.99	11/11/2019	17:10
2	CAL2	1	Water	NA	NA	20.6	7.00	11/11/2019	17:11
3	CAL3	1	Water	NA	NA	20.6	10.03	11/11/2019	17:12
4	ICV	1	Water	NA	NA	20.5	7.01	11/11/2019	17:15
5	CCV1	1	Water	NA	NA	20.5	2.02	11/11/2019	17:16
6	K5670-04	1	Solid	20.03	20	20.9	7.74	11/11/2019	17:19
7	K5670-04DUP	1	Solid	20.04	20	20.8	7.75	11/11/2019	17:20
8	K5788-06	1	Solid	20.03	20	20.2	8.60	11/11/2019	17:21
9	K5789-04	1	Solid	20.04	20	20.4	7.73	11/11/2019	17:22
10	K5789-08	1	Solid	20.03	20	20.1	6.03	11/11/2019	17:24
11	K5789-12	1	Solid	20.05	20	20.4	5.57	11/11/2019	17:25
12	CCV2	1	Water	NA	NA	20.4	12.02	11/11/2019	17:30

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS K5770

WorkList ID : 133077

Department : Wet-Chemistry

Date : 11-11-2019 7:25:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/14/2019	Water	K5762-01-F	TSS	Cool 4 deg C	CAMP02	H61	HP-640	11/07/2019	SM2540 D
11/12/2019	Water	K5764-01 ✓	TSS	Cool 4 deg C	HDR103	H43	110519-JETT-E-D-B	11/05/2019	SM2540 D
11/12/2019	Water	K5764-02 ✓	TSS	Cool 4 deg C	HDR103	H43	110519-JETT-E-D-M	11/05/2019	SM2540 D
11/12/2019	Water	K5764-03 ✓	TSS	Cool 4 deg C	HDR103	H43	110519-JETT-E-D-S	11/05/2019	SM2540 D
11/12/2019	Water	K5764-04 ✓	TSS	Cool 4 deg C	HDR103	H43	110519-JETT-E-U-B	11/05/2019	SM2540 D
11/12/2019	Water	K5764-05 ✓	TSS	Cool 4 deg C	HDR103	H43	110519-JETT-E-U-M	11/05/2019	SM2540 D
11/12/2019	Water	K5764-06 ✓	TSS	Cool 4 deg C	HDR103	H43	110519-JETT-E-U-S	11/05/2019	SM2540 D
11/12/2019	Water	K5764-07 ✓	TSS	Cool 4 deg C	HDR103	H43	110519-SIDI-E-D-B	11/05/2019	SM2540 D
11/12/2019	Water	K5764-08 ✓	TSS	Cool 4 deg C	HDR103	H43	110519-SIDI-E-D-M	11/05/2019	SM2540 D
11/12/2019	Water	K5764-09 ✓	TSS	Cool 4 deg C	HDR103	H43	110519-SIDI-E-D-S	11/05/2019	SM2540 D
11/12/2019	Water	K5764-10 ✓	TSS	Cool 4 deg C	HDR103	H43	110519-SIDI-E-U-B	11/05/2019	SM2540 D
11/12/2019	Water	K5764-11 ✓	TSS	Cool 4 deg C	HDR103	H43	110519-SIDI-E-U-M	11/05/2019	SM2540 D
11/12/2019	Water	K5764-12 ✓	TSS	Cool 4 deg C	HDR103	H43	110519-SIDI-E-U-S	11/05/2019	SM2540 D
11/12/2019	Water	K5764-13 ✓	TSS	Cool 4 deg C	HDR103	H43	110519-DBRI-E-D-B	11/05/2019	SM2540 D
11/12/2019	Water	K5764-14 ✓	TSS	Cool 4 deg C	HDR103	H43	110519-DBRI-E-D-M	11/05/2019	SM2540 D
11/12/2019	Water	K5764-15 ✓	TSS	Cool 4 deg C	HDR103	H43	110519-DBRI-E-D-S	11/05/2019	SM2540 D
11/12/2019	Water	K5764-16 ✓	TSS	Cool 4 deg C	HDR103	H43	110519-DBRI-E-U-B	11/05/2019	SM2540 D
11/12/2019	Water	K5764-17 ✓	TSS	Cool 4 deg C	HDR103	H43	110519-DBRI-E-U-M	11/05/2019	SM2540 D
11/12/2019	Water	K5764-18 ✓	TSS	Cool 4 deg C	HDR103	H43	110519-DBRI-E-U-S	11/05/2019	SM2540 D
11/15/2019	Water	K5766-01 ✓	TSS	Cool 4 deg C	HDR103	H43	110819-DBRI-F-D-B	11/08/2019	SM2540 D
11/15/2019	Water	K5766-02 ✓	TSS	Cool 4 deg C	HDR103	H43	110819-DBRI-F-D-S	11/08/2019	SM2540 D
11/15/2019	Water	K5766-03 ✓	TSS	Cool 4 deg C	HDR103	H43	110819-DBRI-F-U-B	11/08/2019	SM2540 D

Date/Time 11-11-19 10:00

Received by: JP

Relinquished by: CP

Date/Time 11-11-19 16:00

Received by: CP

Relinquished by: JP

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS K5770 WorkList ID : 133077 Department : Wet-Chemistry Date : 11-09-2019 10:17:19

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/15/2019	Water	K5766-04 ✓	TSS	Cool 4 deg C	HDR103	H43	110819-DBRI-F-U-S	11/08/2019	SM2540 D
11/14/2019	Water	K5767-01 ✓	TSS	Cool 4 deg C	HDR103	H43	110719-DBRI-E-D-B	11/07/2019	SM2540 D
11/14/2019	Water	K5767-02 ✓	TSS	Cool 4 deg C	HDR103	H43	110719-DBRI-E-D-M	11/07/2019	SM2540 D
11/14/2019	Water	K5767-03 ✓	TSS	Cool 4 deg C	HDR103	H43	110719-DBRI-E-D-S	11/07/2019	SM2540 D
11/14/2019	Water	K5767-04 ✓	TSS	Cool 4 deg C	HDR103	H43	110719-DBRI-E-U-B	11/07/2019	SM2540 D
11/14/2019	Water	K5767-05 ✓	TSS	Cool 4 deg C	HDR103	H43	110719-DBRI-E-U-M	11/07/2019	SM2540 D
11/14/2019	Water	K5767-06 ✓	TSS	Cool 4 deg C	HDR103	H43	110719-DBRI-E-U-S	11/07/2019	SM2540 D
11/13/2019	Water	K5768-01 ✓	TSS	Cool 4 deg C	HDR103	H43	110619-DBRI-E-D-B	11/06/2019	SM2540 D
11/13/2019	Water	K5768-02 ✓	TSS	Cool 4 deg C	HDR103	H43	110619-DBRI-E-D-M	11/06/2019	SM2540 D
11/13/2019	Water	K5768-03 ✓	TSS	Cool 4 deg C	HDR103	H43	110619-DBRI-E-D-S	11/06/2019	SM2540 D
11/13/2019	Water	K5768-04 ✓	TSS	Cool 4 deg C	HDR103	H43	110619-DBRI-E-U-B	11/06/2019	SM2540 D
11/13/2019	Water	K5768-05 ✓	TSS	Cool 4 deg C	HDR103	H43	110619-DBRI-E-U-M	11/06/2019	SM2540 D
11/13/2019	Water	K5768-06 ✓	TSS	Cool 4 deg C	HDR103	H43	110619-DBRI-E-U-S	11/06/2019	SM2540 D
11/14/2019	Water	K5770-01 ✓	TSS	Cool 4 deg C	HOLL01	H51	EFFLUENT	11/07/2019	SM2540 D
11/14/2019	Water	K5770-04 ✓	TSS	Cool 4 deg C	HOLL01	H51	AER-TANK-1	11/07/2019	SM2540 D

Date/Time 11-11-19 10:00
 Received by: Jo
 Relinquished by: [Signature]

Date/Time 11-11-19 10:00
 Received by: SC
 Relinquished by: [Signature]