



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: Non-Polar Material
Run Number: LB106212
Analysis Date: 11/12/2019
BalanceID: WC SC-3
OvenID: EXT OVEN-3

ANALYST: JIGNESH
REVIEWED BY: apatel
Extraction Date: 11/12/2019
Extraction IN Time: 10:25
Extraction OUT Time: 11:10

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB106212BL	LB106212BL	WATER	1.3	1000	100	2.1657	2.1657	3.02	2.1657	2.1657	0.0000	0
2	LB106212BS	LB106212BS	WATER	1.3	1000	100	2.8974	2.8974	3.01	2.9162	2.9162	0.0188	18.8
3	LB106212BSD	LB106212BSD	WATER	1.3	1000	100	2.9601	2.9601	3.03	2.9791	2.9791	0.0190	19
4	K5787-01	DSN-001-SEWER-OUTLET-0	WATER	1.6	1000	100	2.9446	2.9446	3.04	2.9480	2.9480	0.0034	3.4
5	K5787-02	DSN-002-SEWER-OUTLET-0	WATER	1.6	1000	100	2.9409	2.9409	3.05	2.9447	2.9447	0.0038	3.8



QC Batch# LB106212 Test: Non-Polar Mate Analysis Date: 11/12/2019

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W2612
pH Paper 0-14	M4212
Sodium Sulfate	EP1995
1:1 HCL	WP79191
Silica Gel	W2546
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP79192
LCSWD	10.00 ML	WP79193
MS/MSD	NA	NA

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance:	<u>0.0019</u>	In OVEN TEMP1 :	<u>70 °C</u>	Dessicator Time In1	<u>12:26</u>
1.0000 gram Balance:	<u>1.0005</u>	In Time1:	<u>11:25</u>		
Bal Check Time:	<u>10:35</u>	Out OVEN TEMP1:	<u>70 °C</u>	Dessicator Time Out1	<u>12:50</u>
		Out Time1	<u>12:25</u>		

After Analysis

0.0020 gram Balance:	<u>0.0021</u>	In OVEN TEMP2 :	<u>71 °C</u>	Dessicator Time In2	<u>14:01</u>
1.0000 gram Balance:	<u>1.0003</u>	In Time2:	<u>13:30</u>		
Bal Check Time:	<u>15:00</u>	Out OVEN TEMP2:	<u>71 °C</u>	Dessicator Time Out2	<u>14:30</u>
		Out Time2:	<u>14:00</u>		

WORKLIST(Hardcopy Internal Chain)

2-B106212

WorkList Name : NON POPLAR K5178

WorkList ID : 133156

Department : Wet-Chemistry

Date : 11-12-2019 10:19:21

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/25/2019	Water	K5787-01	Non-Polar Material	Conc H2SO4 10	TRIS02	H43	DSN-001-SEWER-OUTLET-01	11/11/2019	1664A
11/25/2019	Water	K5787-02	Non-Polar Material	Conc H2SO4 10	TRIS02	H43	DSN-002-SEWER-OUTLET-01	11/11/2019	1664A

Date/Time 11-12-19 10:20

Received by: JN

Relinquished by: JN

Date/Time 11-12-19 11:45

Received by: JN

Relinquished by: JN