



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB106213

Slope : 98.2

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	11/11/2019	18:10
2	CAL2	1	Water	NA	NA	20.6	7.00	11/11/2019	18:11
3	CAL3	1	Water	NA	NA	20.6	10.03	11/11/2019	18:12
4	ICV	1	Water	NA	NA	20.5	7.00	11/11/2019	18:15
5	CCV1	1	Water	NA	NA	20.6	2.02	11/11/2019	18:16
6	K5721-01	1	Water	NA	NA	17.3	7.06	11/11/2019	18:19
7	K5721-01DUP	1	Water	NA	NA	17.4	7.07	11/11/2019	18:20
8	CCV2	1	Water	NA	NA	20.5	12.02	11/11/2019	18:21

WORKLIST(Hardcopy Internal Chain)

106209
106213

WorkList Name : PH K5721 WorkList ID : 133132 Department : Wet-Chemistry Date : 11-11-2019 17:57:21

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/11/2019	Water	K5721-01-1	pH	Cool 4 deg C	WAYN02	H42	WASTEWATER	11/11/2019	9040C

Date/Time 11-11-19 18:00
Received by: JA
Relinquished by: CD

Date/Time 11-11-19 18:35
Received by: CD
Relinquished by: JA