



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB106214

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	11/12/2019	11:00
2	CAL2	1	Water	NA	NA	21.3	7.01	11/12/2019	11:01
3	CAL3	1	Water	NA	NA	21.3	10.03	11/12/2019	11:02
4	ICV	1	Water	NA	NA	21.3	7.00	11/12/2019	11:04
5	CCV1	1	Water	NA	NA	21.3	2.02	11/12/2019	11:05
6	K5791-01	1	Solid	20.03	20	22.2	9.89	11/12/2019	11:06
7	K5791-01DUP	1	Solid	20.04	20	22.3	9.90	11/12/2019	11:07
8	CCV2	1	Water	NA	NA	21.4	12.02	11/12/2019	11:10

2B106214

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K5791 WorkList ID : 133158 Department : Wet-Chemistry Date : 11-12-2019 10:48:03

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/11/2019	Solid	K5791-01-P	pH	Cool 4 deg C	BRO004	H51	FINES-NORTH-BERGEN-1	11/11/2019	9045D

Date/Time 11-12-19 10:50
Received by: JD
Relinquished by: JD

Date/Time 11-12-19 11:50
Received by: cd
Relinquished by: JD