


LB106217
INSTRUMENT ID :- CV1

ISM02.4



Linear

A= 0.0000e+000

B= 2.1969e-004

C= -1.5638e-004

Rho= 0.9995967

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.00	0.000	0.043	0.043	197	0.000	197				
0.05	0.050									
0.20	0.200	0.239	0.039	1089	0.0 %	1089				
2.50	2.500	2.546	0.046	11589	0.0 %	11589				
5.00	5.000	4.838	-0.162	22021	0.0 %	22021				
7.50	7.500	7.388	-0.112	33629	0.0 %	33629				
10.0	10.000	10.146	0.146	46186	0.0 %	46186				

1.0

+20
+2%
-3%
-4%
1%.

LB106271
INSTRUMENT ID CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0 S0		197	-		0 ISM02.4	PPB	11/12/2019 13:24	Std
0.2 S0.2		1089	-		0.2 ISM02.4	PPB	11/12/2019 13:26	Std
2.5 S2.5		11589	-		2.5 ISM02.4	PPB	11/12/2019 13:28	Std
5 S5.0		22021	-		5 ISM02.4	PPB	11/12/2019 13:31	Std
7.5 S7.5		33629	-		7.5 ISM02.4	PPB	11/12/2019 13:33	Std
10 S10.0		46186	-		10 ISM02.4	PPB	11/12/2019 13:35	Std
ICV ICB096	ICV ICB096	16605	3.6478	-	ISM02.4	PPB	11/12/2019 13:38	SMPL
ICB	ICB	151	0.033	-	ISM02.4	PPB	11/12/2019 13:41	SMPL
CCV029	CCV029	22909	5.0327	-	ISM02.4	PPB	11/12/2019 13:43	SMPL
CCB029	CCB029	134	0.0293	-	ISM02.4	PPB	11/12/2019 13:45	SMPL
PB124709BL	PBW709	166	0.0363	-	ISM02.4	PPB	11/12/2019 13:47	SMPL
K5744-01	MF2J00	256	0.0561	-	ISM02.4	PPB	11/12/2019 13:50	SMPL
K5744-02	MF2J00D	254	0.0556	-	ISM02.4	PPB	11/12/2019 13:52	SMPL
K5744-03	MF2J00S	5290	1.162	-	ISM02.4	PPB	11/12/2019 13:54	SMPL
K5744-04	MF2J02	229	0.0502	-	ISM02.4	PPB	11/12/2019 13:56	SMPL
K5744-05	MF2J04	404	0.0886	-	ISM02.4	PPB	11/12/2019 13:59	SMPL
K5744-06	MF2J06	172	0.0376	-	ISM02.4	PPB	11/12/2019 14:01	SMPL
K5744-07	MF2J08	168	0.0368	-	ISM02.4	PPB	11/12/2019 14:03	SMPL
K5744-08	MF2J10	170	0.0372	-	ISM02.4	PPB	11/12/2019 14:06	SMPL
K5744-09	MF2J20	142	0.031	-	ISM02.4	PPB	11/12/2019 14:08	SMPL
K5744-10	MF2J30	169	0.037	-	ISM02.4	PPB	11/12/2019 14:10	SMPL
K5744-11	MF2J14	153	0.0335	-	ISM02.4	PPB	11/12/2019 14:12	SMPL
K5744-12	MF2J16	146	0.0319	-	ISM02.4	PPB	11/12/2019 14:15	SMPL
K5744-13	MF2J18	155	0.0339	-	ISM02.4	PPB	11/12/2019 14:17	SMPL
K5744-14	MF2J22	144	0.0315	-	ISM02.4	PPB	11/12/2019 14:19	SMPL
K5744-15	MF2J24	138	0.0302	-	ISM02.4	PPB	11/12/2019 14:21	SMPL
K5744-16	MF2J26	158	0.0346	-	ISM02.4	PPB	11/12/2019 14:24	SMPL
K5744-17	MF2J28	191	0.0418	-	ISM02.4	PPB	11/12/2019 14:26	SMPL
K5744-18	MF2J34	135	0.0295	-	ISM02.4	PPB	11/12/2019 14:28	SMPL
K5744-19	MF2J32	239	0.0523	-	ISM02.4	PPB	11/12/2019 14:31	SMPL
CCV030	CCV030	23096	5.0738	-	ISM02.4	PPB	11/12/2019 14:33	SMPL
CCB030	CCB030	147	0.0321	-	ISM02.4	PPB	11/12/2019 14:35	SMPL
K5744-20	MF2J36	245	0.0537	-	ISM02.4	PPB	11/12/2019 14:37	SMPL
K5744-21	MF2J38	219	0.048	-	ISM02.4	PPB	11/12/2019 14:40	SMPL
PB124710BL	PBW710	169	0.037	-	ISM02.4	PPB	11/12/2019 14:42	SMPL
K5745-01	MF2J01	209	0.0458	-	ISM02.4	PPB	11/12/2019 14:44	SMPL
K5745-02	MF2J01D	201	0.044	-	ISM02.4	PPB	11/12/2019 14:46	SMPL
K5745-03	MF2J01S	5025	1.1038	-	ISM02.4	PPB	11/12/2019 14:49	SMPL
K5745-04	MF2J03	191	0.0418	-	ISM02.4	PPB	11/12/2019 14:51	SMPL
K5745-05	MF2J05	216	0.0473	-	ISM02.4	PPB	11/12/2019 14:53	SMPL
K5745-06	MF2J07	185	0.0405	-	ISM02.4	PPB	11/12/2019 14:55	SMPL
K5745-07	MF2J09	156	0.0341	-	ISM02.4	PPB	11/12/2019 14:58	SMPL
K5745-08	MF2J11	150	0.0328	-	ISM02.4	PPB	11/12/2019 15:00	SMPL
K5745-09	MF2J31	178	0.0389	-	ISM02.4	PPB	11/12/2019 15:02	SMPL
K5745-10	MF2J15	159	0.0348	-	ISM02.4	PPB	11/12/2019 15:05	SMPL
K5745-11	MF2J17	150	0.0328	-	ISM02.4	PPB	11/12/2019 15:07	SMPL

ICV
11/12/19

LB106271
INSTRUMENT ID CV1

K5745-12	MF2J19	162	0.0354 -	ISM02.4	PPB	11/12/2019 15:09 SMPL
K5745-13	MF2J23	151	0.033 -	ISM02.4	PPB	11/12/2019 15:11 SMPL
K5745-14	MF2J25	159	0.0348 -	ISM02.4	PPB	11/12/2019 15:14 SMPL
K5745-15	MF2J27	154	0.0337 -	ISM02.4	PPB	11/12/2019 15:16 SMPL
K5745-16	MF2J29	210	0.046 -	ISM02.4	PPB	11/12/2019 15:18 SMPL
K5745-17	MF2J35	156	0.0341 -	ISM02.4	PPB	11/12/2019 15:20 SMPL
CCV031	CCV031	22824	5.0141 -	ISM02.4	PPB	11/12/2019 15:23 SMPL
CCB031	CCB031	143	0.0313 -	ISM02.4	PPB	11/12/2019 15:25 SMPL
K5745-18	MF2J37	153	0.0335 -	ISM02.4	PPB	11/12/2019 15:27 SMPL
K5745-19	MF2J39	145	0.0317 -	ISM02.4	PPB	11/12/2019 15:30 SMPL
PB124711BL	PBW711	156	0.0341 -	ISM02.4	PPB	11/12/2019 15:32 SMPL
K5653-01	MBFPQ5	7209	1.5836 -	ISM02.4	PPB	11/12/2019 15:34 SMPL
K5653-02	MBFPQ7	578	0.1268 -	ISM02.4	PPB	11/12/2019 15:36 SMPL
K5653-03	MBFPQ7D	647	0.142 -	ISM02.4	PPB	11/12/2019 15:39 SMPL
K5653-04	MBFPQ7S	5728	1.2582 -	ISM02.4	PPB	11/12/2019 15:41 SMPL
K5653-05	MBFPQ8	804	0.1765 -	ISM02.4	PPB	11/12/2019 15:43 SMPL
K5653-06	MBFPQ9	141	0.0308 -	ISM02.4	PPB	11/12/2019 15:45 SMPL
K5653-07	MBFPQ2	103	0.0225 -	ISM02.4	PPB	11/12/2019 15:48 SMPL
K5653-08	MBFPQ6	115	0.0251 -	ISM02.4	PPB	11/12/2019 15:50 SMPL
K5653-09	MBFPR0	109	0.0238 -	ISM02.4	PPB	11/12/2019 15:52 SMPL
K5653-10	MBFPQ3	2447	0.5374 -	ISM02.4	PPB	11/12/2019 15:55 SMPL
K5653-11	MBFPQ4	107	0.0234 -	ISM02.4	PPB	11/12/2019 15:57 SMPL
K5653-12	MBFPR1	124	0.0271 -	ISM02.4	PPB	11/12/2019 15:59 SMPL
PB124712BL	PBW712	191	0.0418 -	ISM02.4	PPB	11/12/2019 16:01 SMPL
K5638-07	MX2J57	33923	7.4524 -	ISM02.4	PPB	11/12/2019 16:04 SMPL
K5638-08	MX2J57D	33526	7.3652 -	ISM02.4	PPB	11/12/2019 16:06 SMPL
K5638-09	MX2J57S	38462	8.4496 -	ISM02.4	PPB	11/12/2019 16:08 SMPL
CCV032	CCV032	22360	4.9121 -	ISM02.4	PPB	11/12/2019 16:10 SMPL
CCB032	CCB032	14	0.0029 -	ISM02.4	PPB	11/12/2019 16:13 SMPL