



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB106223

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	3.99	11/12/2019	13:00
2	CAL2	1	Water	NA	NA	20.3	7.00	11/12/2019	13:01
3	CAL3	1	Water	NA	NA	20.3	10.03	11/12/2019	13:03
4	ICV	1	Water	NA	NA	20.2	7.01	11/12/2019	13:05
5	CCV1	1	Water	NA	NA	20.2	2.02	11/12/2019	13:06
6	K5795-01	1	Water	NA	NA	18.3	5.12	11/12/2019	13:07
7	K5795-01DUP	1	Water	NA	NA	18.4	5.13	11/12/2019	13:09
8	CCV2	1	Water	NA	NA	20.2	12.02	11/12/2019	13:10

✓ 87106223

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K5795 WorkList ID : 133163 Department : Wet-Chemistry Date : 11-12-2019 12:45:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/07/2019	Water	K5795-01 - <i>PH</i>	pH	Cool 4 deg C	STER06	H51	POOL-WATER	11/07/2019	9040C

Date/Time 11-12-19 12:50
Received by: *Ja*
Relinquished by: *LS*

Date/Time 11-12-19 13:35
Received by: *CS*
Relinquished by: *Ja*