


LB 106 244

INSTRUMENT ID - CUI

7471B



Linear

A= 0.0000e+000

B= 1.8440e-004

C= 2.3836e-002

Rho= 0.9995780

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.031	0.031	40	0.000	40				
0.2	0.200	0.252	0.052	1240	0.0 %	1240				
2.5	2.500	2.553	0.053	13713	0.0 %	13713				
5.0	5.000	4.768	-0.232	25726	0.0 %	25726				
7.5	7.500	7.504	0.004	40565	0.0 %	40565				
10.0	10.000	10.092	0.092	54600	0.0 %	54600				

LB106244 INSTRUMENT ID CV1

Method: 7471B		Operator: Admin		Date of Analysis: 13 Nov 2019 07:37:12					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	40	-	0.0000	7471B	PPB	13 Nov 2019 07:38:36	S	Std
0.2 - 1	50.2	1240	-	0.2000	7471B	PPB	13 Nov 2019 07:40:54	S	Std
2.5 - 1	52.5	13713	-	2.5000	7471B	PPB	13 Nov 2019 07:43:11	S	Std
5.0 - 1	55.0	25726	-	5.0000	7471B	PPB	13 Nov 2019 07:45:29	S	Std
7.5 - 1	57.5	40565	-	7.5000	7471B	PPB	13 Nov 2019 07:47:46	S	Std
10.0 - 1	510.0	54600	-	10.0000	7471B	PPB	13 Nov 2019 07:50:04	S	Std
ICV97 - 1	ICV97	19905	3.6943	-	7471B	PPB	13 Nov 2019 07:54:18	U	SMPL
ICB97 - 1	ICB97	139	0.0495	-	7471B	PPB	13 Nov 2019 07:56:34	U	SMPL
CCV33 - 1	CCV33	25549	4.7351	-	7471B	PPB	13 Nov 2019 07:58:51	U	SMPL
CCB33 - 1	CCB33	73	0.0373	-	7471B	PPB	13 Nov 2019 08:01:06	U	SMPL
CRA - 1	CRA	1264	0.2569	-	7471B	PPB	13 Nov 2019 08:03:23	U	SMPL
HighStd - 1	HighStd	54093	9.9986	-	7471B	PPB	13 Nov 2019 08:05:38	U	SMPL
ChkStd - 1	ChkStd	37465	6.9324	-	7471B	PPB	13 Nov 2019 08:07:54	U	SMPL
PB124743BL - 1	PBS	-52	0.0142	-	7471B	PPB	13 Nov 2019 08:10:12	U	SMPL
PB124743BS - 1	LCSS	18847	3.4992	-	7471B	PPB	13 Nov 2019 08:12:29	U	SMPL
K5670-05 - 1	OK-03-111219	1817	0.3589	-	7471B	PPB	13 Nov 2019 08:14:46	U	SMPL
K5791-01 - 1	FINES-NORTH-BERGEN-1	29485	5.4609	-	7471B	PPB	13 Nov 2019 08:17:04	U	SMPL
K5791-01DUP - 1	FINES-NORTH-BERGEN-1 DUP	34537	6.3925	-	7471B	PPB	13 Nov 2019 08:19:21	U	SMPL
K5791-01MS - 1	FINES-NORTH-BERGEN-1 MS	53209	9.8356	-	7471B	PPB	13 Nov 2019 08:21:38	U	SMPL
K5791-01MSD - 1	FINES-NORTH-BERGEN-1 MSD	52000	9.6126	-	7471B	PPB	13 Nov 2019 08:23:54	U	SMPL
CCV34 - 1	CCV34	25604	4.7452	-	7471B	PPB	13 Nov 2019 08:34:41	U	SMPL
CCB34 - 1	CCB34	168	0.0548	-	7471B	PPB	13 Nov 2019 08:38:50	U	SMPL
K5791-01LX5 - 1		5645	1.0648	-	7471B	PPB	13 Nov 2019 08:41:06	U	SMPL
K5791-01A - 1		53125	9.8201	-	7471B	PPB	13 Nov 2019 08:43:21	U	SMPL
CCV35 - 1	CCV35	25740	4.7703	-	7471B	PPB	13 Nov 2019 08:47:55	U	SMPL
CCB35 - 1	CCB35	53	0.0336	-	7471B	PPB	13 Nov 2019 08:50:12	U	SMPL