

LB106246


INSTRUMENT ID:- CUI

7470A

Linear

 μ Abs.:

41564



A= 0.0000e+000

B= 2.4078e-004

C= -1.0158e-001

Rho= 0.9996962

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.095	-0.095	29	0.000	29				
0.2	0.200	0.148	-0.052	1036	0.0 %	1036				
2.5	2.500	2.661	0.161	11475	0.0 %	11475				
5.0	5.000	5.039	0.039	21351	0.0 %	21351				
7.5	7.500	7.540	0.040	31736	0.0 %	31736				
10.0	10.000	9.906	-0.094	41564	0.0 %	41564				

LB106246 INSTRUMENT ID CV1

Method: 7470A

Operator: Admin

Date of Analysis: 13 Nov 2019 09:24:47

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	29	-	0.0000	7470A	PPB	13 Nov 2019 09:33:05	S	Std
0.2 - 1	50.2	1036	-	0.2000	7470A	PPB	13 Nov 2019 09:35:14	S	Std
2.5 - 1	52.5	11475	-	2.5000	7470A	PPB	13 Nov 2019 09:37:22	S	Std
5.0 - 1	55.0	21351	-	5.0000	7470A	PPB	13 Nov 2019 09:39:30	S	Std
7.5 - 1	57.5	31736	-	7.5000	7470A	PPB	13 Nov 2019 09:41:38	S	Std
10.0 - 1	50.0	41564	-	10.0000	7470A	PPB	13 Nov 2019 09:43:46	S	Std
ICV98 - 1	ICV98	16955	3.9808	-	7470A	PPB	13 Nov 2019 10:01:28	U	SMPL
ICB98 - 1	ICB98	25	-0.0956	-	7470A	PPB	13 Nov 2019 10:03:34	U	SMPL
CCV36 - 1	CCV36	21161	4.9936	-	7470A	PPB	13 Nov 2019 10:05:40	U	SMPL
CCB36 - 1	CCB36	23	-0.0960	-	7470A	PPB	13 Nov 2019 10:07:46	U	SMPL
CRA - 1	CRA	1067	0.1553	-	7470A	PPB	13 Nov 2019 10:09:52	U	SMPL
HighStd - 1	HighStd	41975	10.0052	-	7470A	PPB	13 Nov 2019 10:11:58	U	SMPL
ChkStd - 1	ChkStd	29280	6.9485	-	7470A	PPB	13 Nov 2019 10:14:06	U	SMPL
PB124738BL - 1	PBW	123	-0.0720	-	7470A	PPB	13 Nov 2019 10:16:14	U	SMPL
PB124738BS - 1	LCSW	15759	3.6929	-	7470A	PPB	13 Nov 2019 10:18:21	U	SMPL
K5670-04 - 1	OK-02-111119	-10	-0.1040	-	7470A	PPB	13 Nov 2019 10:20:29	U	SMPL
K5785-02 - 1	72-12002-VNJ-221	23	-0.0960	-	7470A	PPB	13 Nov 2019 10:22:36	U	SMPL
K5788-07 - 1	WC-1	36	-0.0929	-	7470A	PPB	13 Nov 2019 10:24:44	U	SMPL
K5789-04 - 1	TP-17-S	-19	-0.1062	-	7470A	PPB	13 Nov 2019 10:26:53	U	SMPL
K5789-08 - 1	TP-16B-S	49	-0.0898	-	7470A	PPB	13 Nov 2019 10:29:00	U	SMPL
CCV37 - 1	CCV37	21516	5.0790	-	7470A	PPB	13 Nov 2019 10:31:06	U	SMPL
CCB37 - 1	CCB37	26	-0.0953	-	7470A	PPB	13 Nov 2019 10:33:12	U	SMPL
K5789-12 - 1	TP-16B-D	34	-0.0934	-	7470A	PPB	13 Nov 2019 10:35:19	U	SMPL
K5789-12DUP - 1	TP-16B-DDUP	25	-0.0956	-	7470A	PPB	13 Nov 2019 10:37:26	U	SMPL
K5789-12MS - 1	TP-16B-DMS	16425	3.8532	-	7470A	PPB	13 Nov 2019 10:39:33	U	SMPL
K5789-12MSD - 1	TP-16B-DMSD	16521	3.8764	-	7470A	PPB	13 Nov 2019 10:41:40	U	SMPL
K5789-12LX5 - 1		-10	-0.1040	-	7470A	PPB	13 Nov 2019 10:43:47	U	SMPL
K5489-12A - 1 K5184-1	12A	16606	3.8968	-	7470A	PPB	13 Nov 2019 10:45:54	U	SMPL
PB124738TB - 1	PB124707TB	64	-0.0862	-	7470A	PPB	13 Nov 2019 10:48:02	U	SMPL
CCV38 - 1	CCV38	21276	5.0213	-	7470A	PPB	13 Nov 2019 10:50:10	U	SMPL
CCB38 - 1	CCB38	-1	-0.1018	-	7470A	PPB	13 Nov 2019 10:52:17	U	SMPL

CT
11/13/19