


LB106248
INSTRUMENT ID :- CV1

ISM02.4



Linear

A= 0.0000e+000

B= 2.2896e-004 slope

C= -5.0196e-002 y-intercept

Rho= 0.9993000

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	D.V.
0.00	0.000	-0.022	-0.022	124	0.000	124					18.1.
0.05	0.050	0.059	0.009	477	0.0 %	477					11.1.
0.20	0.200	0.223	0.023	1195	0.0 %	1195					3.1.
2.50	2.500	2.584	0.084	11504	0.0 %	11504					-5.1.
5.00	5.000	4.759	-0.241	21005	0.0 %	21005					3.1.
7.50	7.500	7.744	0.244	34040	0.0 %	34040					0.1.
10.0	10.000	9.903	-0.097	43469	0.0 %	43469					

CT
11/13/19

LB106248
INSTRUMENT ID CV1

Sample ID	Extended I	μ Abs.	Conc.	Std Cor Method	Units	Date	Type
0 S0		124	-	0 ISM02.4	PPB	11/13/2019 11:40	Std
0.05 S0.05		477	-	0.05 ISM02.4	PPB	11/13/2019 11:45	Std
0.2 S0.2		1195	-	0.2 ISM02.4	PPB	11/13/2019 11:47	Std
2.5 S2.5		11504	-	2.5 ISM02.4	PPB	11/13/2019 11:49	Std
5 S5		21005	-	5 ISM02.4	PPB	11/13/2019 11:52	Std
7.5 S7.5		34040	-	7.5 ISM02.4	PPB	11/13/2019 11:54	Std
10 S10		43469	-	10 ISM02.4	PPB	11/13/2019 12:09	Std
ICV	ICV	19029	4.3068 -	ISM02.4	PPB	11/13/2019 12:12	SMPL
ICB	ICB	103	-0.0266 -	ISM02.4	PPB	11/13/2019 12:15	SMPL
CCV039	CCV039	21038	4.7667 -	ISM02.4	PPB	11/13/2019 12:17	SMPL
CCB039	CCB039	105	-0.0262 -	ISM02.4	PPB	11/13/2019 12:19	SMPL
PB124711E	PBW711	118	-0.0232 -	ISM02.4	PPB	11/13/2019 12:21	SMPL
K5653-01	MBFPQ5	7588	1.6872 -	ISM02.4	PPB	11/13/2019 12:24	SMPL
K5653-02	MBFPQ7	573	0.081 -	ISM02.4	PPB	11/13/2019 12:26	SMPL
K5653-03	MBFPQ7D	642	0.0968 -	ISM02.4	PPB	11/13/2019 12:28	SMPL
K5653-04	MBFPQ7S	5829	1.2844 -	ISM02.4	PPB	11/13/2019 12:30	SMPL
K5653-05	MBFPQ8	803	0.1337 -	ISM02.4	PPB	11/13/2019 12:33	SMPL
K5653-06	MBFPQ9	149	-0.0161 -	ISM02.4	PPB	11/13/2019 12:35	SMPL
K5653-07	MBFPQ2	119	-0.0229 -	ISM02.4	PPB	11/13/2019 12:37	SMPL
K5653-08	MBFPQ6	109	-0.0252 -	ISM02.4	PPB	11/13/2019 12:40	SMPL
K5653-09	MBFPRO	129	-0.0207 -	ISM02.4	PPB	11/13/2019 12:43	SMPL
K5653-10	MBFPQ3	2692	0.5662 -	ISM02.4	PPB	11/13/2019 12:48	SMPL
K5653-11	MBFPQ4	105	-0.0262 -	ISM02.4	PPB	11/13/2019 12:51	SMPL
K5653-12	MBFPR1	131	-0.0202 -	ISM02.4	PPB	11/13/2019 12:53	SMPL
CCV040	CCV040	21060	4.7718 -	ISM02.4	PPB	11/13/2019 12:55	SMPL
CCB040	CCB040	107	-0.0257 -	ISM02.4	PPB	11/13/2019 12:57	SMPL