



Analytical Summary Report

Analysis Method: 9041A

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB106263

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467
3.0-5.5 pH paper refill	W1934
Test Paper, PH, Short Range 6.0/8.0	W2350
Test Paper,PH Short Range 9.0/10.0	W1939
0.0-3.0 pH paper refill	W1931
10.0-12.0 pH paper refill	W1940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.0	11/13/2019	15:30
2	CAL2	1	Water	NA	NA	20.5	7.0	11/13/2019	15:31
3	CAL3	1	Water	NA	NA	20.6	10.0	11/13/2019	15:32
4	ICV	1	Water	NA	NA	20.6	7.0	11/13/2019	15:33
5	CCV1	1	Water	NA	NA	20.4	2.0	11/13/2019	15:35
6	K5818-01	1	Solid	20.02	20	20.5	6.0	11/13/2019	15:36
7	K5818-03	1	Solid	20.02	20	20.1	6.6	11/13/2019	15:38
8	K5818-03DUP	1	Solid	20.03	20	20.2	6.8	11/13/2019	15:39
9	CCV2	1	Water	NA	NA	20.4	12.0	11/13/2019	15:40

18106263

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH S K5818 WorkList ID : 133221 Department : Wet-Chemistry Date : 11-13-2019 15:18:49

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/13/2019	Solid	K5818-01 - F	pH	Cool 4 deg C	PSEG01	N62	30525	11/13/2019	9041A
11/13/2019	Solid	K5818-03 - F	pH	Cool 4 deg C	PSEG01	N62	40180	11/13/2019	9041A

Date/Time 11-13-19 15:20
Received by: J
Relinquished by: [Signature]

Date/Time 11-13-19 16:30
Received by: J
Relinquished by: J